

















Calibration Report

Database File: cbl noble lds d17-24d.db
Dataset Pathname: pass3
Dataset Creation: Tue Jun 19 12:10:09 2012 by Log SCH 120430

Gamma Ray Calibration Report

Type / Serial: Probe275dig / 110809

SHOP CALIBRATION Tue Jan 03 10:16:04 2012

	Counts/Sec.	Gain	Offset	Jig	Units
Background	0.0				cps
Calibrator	1.0				cps
		0.6800			GAPI/cps

PRIMARY VERIFICATION

Background	0.0				cps
Calibrator	0.0				cps
Difference				0.0	GAPI

BEFORE SURVEY VERIFICATION

Background	0.0				cps
Calibrator	0.0				cps
Difference				0.0	GAPI

AFTER SURVEY VERIFICATION

Background	0.0				cps
Calibrator	0.0				cps
Difference				0.0	GAPI

Cement Bond Log Calibration Report

Serial Number: 110803
Tool Model: Probe275Digital
Performed: Tue Jun 19 11:37:58 2012

Depth: 538.692 ft
Casing Diameter: 4.5 in

	3' Spacing	5' Spacing	
Signal Zero:	0.9	0	mV
Calibrated Amplitude:	81.1961	81.1961	mV
Reading at Signal Zero:	-0.00556641	-0.00561523	V
Reading in Free Pipe:	2.48123	2.34417	V
Gain:	32.289	34.5548	
Offset:	1.07973	0.194033	