



GAS MEASUREMENT EMISSIONS TESTING LABORATORY

307.856.0866
www.precision-labs.com

Run File C:\Galaxie\data\14_05_09\14050911-04 ONE1.DATA

Method S2_H2SOnly

Operator User1 **Analysis Date** 5/9/2014

Client: Mull Drilling **Date Sampled:** 05/07/2014

Sample Identification: APC Baughman Unit **Purpose:** NI

Unique #: 0 **Pressure:** 38 PSI

Sample Temperature: 90.7 DEG F **Type Sample:** SPOT

Sampled by: Erik Hollaway **County:** Cheyenne

Sample Container: Tedlar Bag (1000 cc) **Analysis Method:** Lab GC-TCD

Sample Notes:

Due to the reactive nature of H₂S, results detected on the laboratory gas chromatograph may be lower than the actual result or results detected on-site.

Component	Result	Units
Hydrogen Sulfide	0.0002	Mole %
Hydrogen Sulfide	2	ppm