

Session Report

1/10/2015

Information Panel

Name ALICIA #12-15H-5N
Start Time 1/10/2015 9:46:11 AM
Stop Time 1/10/2015 10:50:45 AM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Calibration History

<u>Date</u>	<u>Calibration Action</u>	<u>Level</u>	<u>Cal. Model Type</u>	<u>Serial Number</u>	<u>Cert. Due Date</u>
1/10/2015 9:45:58 AM	Calibration	114.0			
1/10/2015 10:50:33 AM	Verification	114.2			

Summary Data Panel

<u>Description</u>	<u>Meter</u>	<u>Value</u>	<u>Description</u>	<u>Meter</u>	<u>Value</u>
Leq	1	63.3 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	77.4 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

ALICIA #12-15H-5N: Logged Data Chart

