

Session Report

1/28/2015

Information Panel

Name Young01N65W287N_12715
Start Time 1/27/2015 6:33:18 PM
Stop Time 1/27/2015 7:20:29 PM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
1/27/2015 6:32:58 PM	Calibration	114.0			
1/27/2015 7:20:51 PM	Verification	114.1			

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	42.7 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	59.4 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

Young01N65W287N_12715: Logged Data Chart

