

Session Report

3/2/2015

Information Panel

Name NGL#C5_030215
Start Time 3/2/2015 1:01:29 PM
Stop Time 3/2/2015 1:59:47 PM
Device Name BGN010013
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
3/2/2015 1:00:56 PM	Calibration	114.1			
3/2/2015 2:00:16 PM	Verification	114.1			

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	49.2 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	60.1 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			

Logged Data Chart

NGL#C5_030215: Logged Data Chart

