

Session Report

3/18/2015

Information Panel

Name S085_BGN010013_18032015_143335
Start Time 3/18/2015 1:00:17 PM
Stop Time 3/18/2015 2:26:13 PM
Device Name BGN010013
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Calibration History

<u>Date</u>	<u>Calibration Action</u>	<u>Level</u>	<u>Cal. Model Type</u>	<u>Serial Number</u>	<u>Cert. Due Date</u>
3/18/2015 1:00:01 PM	Calibration	114.0			
3/18/2015 2:26:40 PM	Verification	114.0			

Summary Data Panel

<u>Description</u>	<u>Meter</u>	<u>Value</u>	<u>Description</u>	<u>Meter</u>	<u>Value</u>
Leq	1	49.5 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	61.2 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			

Logged Data Chart

S085_BGN010013_18032015_143335: Logged Data Chart

