

Session Report

4/7/2015

Information Panel

Name DT Martinez location 3-31-2015
Start Time 3/31/2015 9:52:58 PM
Stop Time 3/31/2015 10:53:06 PM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments GPS of sound study 40.43826 -104.67795

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	54.7 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	75.2 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
3/31/2015 9:49:19 PM	Calibration	114.0			
3/31/2015 10:56:05 PM	Verification	114.1			

Logged Data Chart

DT Martinez location 3-31-2015: Logged Data Chart

