



















Calibration Report
Ordering No. 73800-06
Customer Number 0000-0000
Calibration Interval 12 mo. as of 30-Apr-2015
Service Tag / Calibration Report
Serial Number 0000
Model Number Fluke 719 30G II, 0 psi to 30 psi
Performed by Fluke
Calibration Date 30-Apr-2014
Temperature 21 °C
Humidity Reading 23 %RH
Calibration Facility USA
Ingredients Control Device Calibration Report
Serial Number 0000
Model Number Fluke 719 30G II, 0 psi to 30 psi
Calibration Procedure Number 2010
Calibration Due Date 30-Apr-2015
Device calibration, performed externally
Test In
Calibrated Out
Results
Zero
Span
Zero
Span
Zero
Span
0%
0.000
30.000
0.000
30.000
In Tol
In Tol
10%
3.000
27.000
3.000
27.000
In Tol
In Tol
20%
6.000
24.000
6.000
24.000
In Tol
In Tol
30%
9.000
21.000
9.000
21.000
In Tol
In Tol
40%
12.000
18.000
12.000
18.000
In Tol
In Tol
50%
15.000
15.000
15.000
15.000
In Tol
In Tol
60%
18.000
12.000
18.000
12.000
In Tol
In Tol
70%
21.000
9.000
21.000
9.000
In Tol
In Tol
80%
24.000
6.000
24.000
6.000
In Tol
In Tol
90%
27.000
3.000
27.000
3.000
In Tol
In Tol
100%
30.000
0.000
30.000
0.000
In Tol
In Tol
Internal Reference Calibration, performed Tue May 27 10:01:12 2014
Test In
Calibrated Out
Results
Zero
Span
Zero
Span
Zero
Span
0.0%
0.010
0.005
0.010
0.005
In Tol
In Tol
As Found and As Left Limits: +/- 0.100 % of span
Test In
Calibrated Out
Results
Zero
Span
Zero
Span
Zero
Span
0%
0.000
0.000
0.000
0.000
In Tol
In Tol
10%
3.000
3.000
3.000
3.000
In Tol
In Tol
20%
6.000
6.000
6.000
6.000
In Tol
In Tol
30%
9.000
9.000
9.000
9.000
In Tol
In Tol
40%
12.000
12.000
12.000
12.000
In Tol
In Tol
50%
15.000
15.000
15.000
15.000
In Tol
In Tol
60%
18.000
18.000
18.000
18.000
In Tol
In Tol
70%
21.000
21.000
21.000
21.000
In Tol
In Tol
80%
24.000
24.000
24.000
24.000
In Tol
In Tol
90%
27.000
27.000
27.000
27.000
In Tol
In Tol
100%
30.000
30.000
30.000
30.000
In Tol
In Tol