

Session Report

10/6/2016

Information Panel


Name Anderson100616
Start Time 10/6/2016 3:30:23 PM
Stop Time 10/6/2016 3:50:05 PM
Device Name BGN010009
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	63.9 dB	Weighting	1	A
Exchange Rate	1	3 dB	Bandwidth	1	OFF
Response	1	FAST			
Leq	2	70.8 dB	Weighting	2	C
Exchange Rate	2	3 dB			
Response	2	FAST			

Logged Data Chart

Anderson100616: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
10/6/2016 3:30:08 PM	Calibration	114.0			
10/6/2016 3:50:20 PM	Verification	113.9			