

Session Report

11/15/2016

Information Panel


Name Boyd111516
Start Time 11/15/2016 2:40:12 PM
Stop Time 11/15/2016 3:11:11 PM
Device Name BGN010009
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

| Description | Meter | Value | Description | Meter | Value |
|---------------|-------|---------|-------------|-------|-------|
| Leq | 1 | 47.3 dB | | | |
| Exchange Rate | 1 | 3 dB | Weighting | 1 | A |
| Response | 1 | SLOW | Bandwidth | 1 | OFF |
| Leq | 2 | 62.9 dB | | | |
| Exchange Rate | 2 | 3 dB | Weighting | 2 | C |
| Response | 2 | FAST | | | |

Logged Data Chart

Boyd111516: Logged Data Chart



Calibration History

| Date | Calibration Action | Level | Cal. Model Type | Serial Number | Cert. Due Date |
|-----------------------|--------------------|-------|-----------------|---------------|----------------|
| 11/15/2016 2:40:00 PM | Calibration | 114.0 | | | |
| 11/15/2016 3:11:29 PM | Verification | 113.9 | | | |