

Session Report

12/12/2016

Information Panel


Name	Champlin Location
Start Time	12/12/2016 11:05:52 AM
Stop Time	12/12/2016 12:06:19 PM
Device Name	BGN010010
Model Type	SoundPro DL
Device Firmware Rev	R.13F
Comments	Winds approx 1-3 MPH out of the NW approx GPS 40.04763 -104.015 Temp approx 30 F

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	54.7 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	FAST	Bandwidth	1	OFF
Leq	2	72.4 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

Champlin Location: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
12/12/2016 AM	10:58:14 Calibration	114.0			

12/12/2016 12:16:31 Verification 113.9
PM