

# Session Report

2/14/2017

## Information Panel


Name LandFed021417  
Start Time 2/14/2017 9:40:40 AM  
Stop Time 2/14/2017 10:35:13 AM  
Device Name BGN010009  
Model Type SoundPro DL  
Device Firmware Rev R.13F  
Comments

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	2	60.5 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

## Logged Data Chart

LandFed021417: Logged Data Chart



## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
2/14/2017 9:40:30 AM	Calibration	114.0			
2/14/2017 10:35:39 AM	Verification	113.7			