

# Session Report

2/14/2017

## Information Panel


Name Jacobucci021417  
Start Time 2/14/2017 11:22:42 AM  
Stop Time 2/14/2017 11:51:42 AM  
Device Name BGN010009  
Model Type SoundPro DL  
Device Firmware Rev R.13F  
Comments

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	45.1 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	59.6 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

## Logged Data Chart

Jacobucci021417: Logged Data Chart



## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
2/14/2017 11:22:30 AM	Calibration	114.0			
2/14/2017 11:52:30 AM	Verification	114.0			