

Session Report

4/11/2017

Information Panel


Name SHOOK041117
Start Time 4/11/2017 9:09:34 AM
Stop Time 4/11/2017 10:09:37 AM
Device Name BGN010009
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	62.2 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	82.2 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

SHOOK041117: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
4/11/2017 9:09:03 AM	Calibration	114.0			
4/11/2017 10:10:10 AM	Verification	113.8			