

Session Report

6/6/2017

Information Panel

Name S084_BGN010009_06062017_104633

Start Time 6/6/2017 9:42:28 AM

Stop Time 6/6/2017 10:42:29 AM

Device Name BGN010009

Model Type SoundPro DL

Device Firmware Rev R.13F


Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	56.7 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	71.6 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

S084_BGN010009_06062017_104633: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
6/6/2017 9:39:59 AM	Calibration	114.0			
6/6/2017 10:43:08 AM	Verification	113.8			