

Session Report

9/15/2017

Information Panel


Name S090_BGN010013_15092017_104047
Start Time 9/15/2017 9:34:06 AM
Stop Time 9/15/2017 10:34:07 AM
Device Name BGN010013
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	57.5 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	76.3 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

S090_BGN010013_15092017_104047: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
9/15/2017 9:31:05 AM	Calibration	114.0			
9/15/2017 10:35:42 AM	Verification	113.9			