

Session Report

10/20/2017

Information Panel


Name S102_BGN010013_20102017_112003
Start Time 10/20/2017 10:07:53 AM
Stop Time 10/20/2017 11:11:55 AM
Device Name BGN010013
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	52.5 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	FAST	Bandwidth	1	OFF
Leq	2	67.8 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

S102_BGN010013_20102017_112003: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
10/20/2017 AM	10:06:42 Calibration	114.0			

10/20/2017 11:17:01 Verification 113.9
AM