

HILL RANCH 27-12R API # 05-071-07183 – TIMBER CREEK  
 OGCC OPERATOR NUMBER 10672 - DATE: OCTOBER 23, 2018  
 CORRECTIVE ACTIONS COMPLETED REGARDING COGCC FIR DOCUMENT # 687903214

DATE	CALIBRATED		ORIFICE INSPECTED	PLATE SIZE	GAS SAMPLE TAKEN	COMMENTS
	Y/N	INITIALS				
9-13-13	yes	JS	yes	.750	NO	Cal DP AND SP. VER. RTD
05-20-14	no	JP	no	0.750	yes 0.5590	GAS Analysis J. Petree
9-19-14	yes	AA	yes	.750 → .625	NO	Ver DP SP RTD → .625
05-15-15	no	JP	no	0.625	yes 0.5593	GAS Analysis J. Petree
9-3-15	yes	MM	yes	.625	NO	VER. FRED
05-03-16	no	JP	no	0.625	yes 0.5584	GAS Analysis J. Petree
9-6-16	yes	AD	yes	0.625	NO	verified
05-16-17	no	JP	no	0.625	yes 0.5596	GAS Analysis J. Petree
9-5-17	yes	AD	yes	0.625	NO	Cal RTD
10-9-18	y	JA	change	.750	NO	JA (9/ SP

THE HILL RANCH 27-12R CALIBRATION IS COMPLETE AND THE CALIBRATION REPORT HAS BEEN UPDATED.