

Session Report

4/3/2019

Information Panel

Name S020_BGN010013_03042019_125745

Start Time 4/3/2019 9:36:51 AM

Stop Time 4/3/2019 10:11:58 AM

Device Name BGN010013

Model Type SoundPro DL

Device Firmware Rev R.13H

Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	57.5 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	72.6 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

S020_BGN010013_03042019_125745: Logged Data Chart

