

# Session Report

4/3/2019

## Information Panel


Name S029\_BGN010010\_03042019\_160309  
Start Time 4/2/2019 3:24:12 PM  
Stop Time 4/3/2019 10:35:52 AM  
Device Name BGN010010  
Model Type SoundPro DL  
Device Firmware Rev R.13F  
Comments See photo attachments for weather data.

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	47.1 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	60.7 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

## Logged Data Chart

S029\_BGN010010\_03042019\_160309: Logged Data Chart



## Calibration History

<u>Date</u>	<u>Calibration Action</u>	<u>Level</u>	<u>Cal. Model Type</u>	<u>Serial Number</u>	<u>Cert. Due Date</u>
4/2/2019 3:22:13 PM	Calibration	114.0			
4/3/2019 10:36:37 AM	Verification	114.1			