

October 07, 2019

<sup>1</sup> Cp

<sup>2</sup> Tc

<sup>3</sup> Ss

<sup>4</sup> Cn

<sup>5</sup> Sr

<sup>6</sup> Qc

<sup>7</sup> Gl

<sup>8</sup> Al

<sup>9</sup> Sc

## Laramie Energy - Grand Junction, CO

Sample Delivery Group: L1145786  
Samples Received: 10/03/2019  
Project Number:  
Description: KOBE FLANGE - CO River  
  
Report To: Robert Stockton  
760 Horizon Dr., Ste. 101  
Grand Junction, CO 81506

Entire Report Reviewed By:

*Chris Ward*

Chris Ward  
Project Manager

Results relate only to the items tested or calibrated and are reported as rounded values. This test report shall not be reproduced, except in full, without written approval of the laboratory. Where applicable, sampling conducted by Pace Analytical National is performed per guidance provided in laboratory standard operating procedures ENV-SOP-MTJL-0067 and ENV-SOP-MTJL-0068. Where sampling conducted by the customer, results relate to the accuracy of the information provided, and as the samples are received.





Cp: Cover Page	1
Tc: Table of Contents	2
Ss: Sample Summary	3
Cn: Case Narrative	4
Sr: Sample Results	5
KOBE FLANGE - CO RIVER L1145786-01	5
Qc: Quality Control Summary	6
Gravimetric Analysis by Method 2540 C-2011	6
Wet Chemistry by Method 9056A	7
Volatile Organic Compounds (GC) by Method 8015D/GRO	9
Volatile Organic Compounds (GC/MS) by Method 8260B	10
Semi-Volatile Organic Compounds (GC) by Method 3511/8015	11
Gl: Glossary of Terms	12
Al: Accreditations & Locations	13
Sc: Sample Chain of Custody	14



# SAMPLE SUMMARY

ONE LAB. NATIONWIDE.



KOBE FLANGE - CO RIVER L1145786-01 GW

Collected by  
Matt Kasten

Collected date/time  
10/02/19 08:00

Received date/time  
10/03/19 08:45

Method	Batch	Dilution	Preparation date/time	Analysis date/time	Analyst	Location
Gravimetric Analysis by Method 2540 C-2011	WG1356912	1	10/06/19 10:53	10/06/19 14:57	TH	Mt. Juliet, TN
Wet Chemistry by Method 9056A	WG1356698	5	10/03/19 22:01	10/03/19 22:01	ELN	Mt. Juliet, TN
Volatile Organic Compounds (GC) by Method 8015D/GRO	WG1357774	1	10/05/19 15:37	10/05/19 15:37	JAH	Mt. Juliet, TN
Volatile Organic Compounds (GC/MS) by Method 8260B	WG1357950	1	10/05/19 12:07	10/05/19 12:07	JHH	Mt. Juliet, TN
Semi-Volatile Organic Compounds (GC) by Method 3511/8015	WG1356886	1	10/03/19 20:54	10/06/19 07:04	JN	Mt. Juliet, TN

<sup>1</sup> Cp

<sup>2</sup> Tc

<sup>3</sup> Ss

<sup>4</sup> Cn

<sup>5</sup> Sr

<sup>6</sup> Qc

<sup>7</sup> Gl

<sup>8</sup> Al

<sup>9</sup> Sc

ACCOUNT:

Laramie Energy - Grand Junction, CO

PROJECT:

SDG:

L1145786

DATE/TIME:

10/07/19 15:44

PAGE:

3 of 14



All sample aliquots were received at the correct temperature, in the proper containers, with the appropriate preservatives, and within method specified holding times, unless qualified or notated within the report. Where applicable, all MDL (LOD) and RDL (LOQ) values reported for environmental samples have been corrected for the dilution factor used in the analysis. All Method and Batch Quality Control are within established criteria except where addressed in this case narrative, a non-conformance form or properly qualified within the sample results. By my digital signature below, I affirm to the best of my knowledge, all problems/anomalies observed by the laboratory as having the potential to affect the quality of the data have been identified by the laboratory, and no information or data have been knowingly withheld that would affect the quality of the data.

Chris Ward  
Project Manager

<sup>1</sup> Cp

<sup>2</sup> Tc

<sup>3</sup> Ss

<sup>4</sup> Cn

<sup>5</sup> Sr

<sup>6</sup> Qc

<sup>7</sup> Gl

<sup>8</sup> Al

<sup>9</sup> Sc



## Gravimetric Analysis by Method 2540 C-2011

Analyte	Result	Qualifier	RDL	Dilution	Analysis date / time	Batch
Dissolved Solids	491		10.0	1	10/06/2019 14:57	<a href="#">WG1356912</a>

## Wet Chemistry by Method 9056A

Analyte	Result	Qualifier	RDL	Dilution	Analysis date / time	Batch
Chloride	157		5.00	5	10/03/2019 22:01	<a href="#">WG1356698</a>
Sulfate	102		25.0	5	10/03/2019 22:01	<a href="#">WG1356698</a>

## Volatile Organic Compounds (GC) by Method 8015D/GRO


Analyte	Result	Qualifier	RDL	Dilution	Analysis date / time	Batch
TPH (GC/FID) Low Fraction	ND		0.100	1	10/05/2019 15:37	<a href="#">WG1357774</a>
(S) <i>a,a,a</i> -Trifluorotoluene(FID)	104		78.0-120		10/05/2019 15:37	<a href="#">WG1357774</a>

## Volatile Organic Compounds (GC/MS) by Method 8260B

Analyte	Result	Qualifier	RDL	Dilution	Analysis date / time	Batch
Benzene	ND		0.00100	1	10/05/2019 12:07	<a href="#">WG1357950</a>
Toluene	ND		0.00100	1	10/05/2019 12:07	<a href="#">WG1357950</a>
Ethylbenzene	ND		0.00100	1	10/05/2019 12:07	<a href="#">WG1357950</a>
Total Xylenes	ND		0.00300	1	10/05/2019 12:07	<a href="#">WG1357950</a>
(S) <i>Toluene-d8</i>	101		80.0-120		10/05/2019 12:07	<a href="#">WG1357950</a>
(S) <i>4-Bromofluorobenzene</i>	101		77.0-126		10/05/2019 12:07	<a href="#">WG1357950</a>
(S) <i>1,2-Dichloroethane-d4</i>	109		70.0-130		10/05/2019 12:07	<a href="#">WG1357950</a>

## Semi-Volatile Organic Compounds (GC) by Method 3511/8015

Analyte	Result	Qualifier	RDL	Dilution	Analysis date / time	Batch
TPH (GC/FID) High Fraction	ND		0.100	1	10/06/2019 07:04	<a href="#">WG1356886</a>
(S) <i>o</i> -Terphenyl	71.6		31.0-160		10/06/2019 07:04	<a href="#">WG1356886</a>



Method Blank (MB)

(MB) R3458404-1 10/06/19 14:57

	MB Result	MB Qualifier	MB MDL	MB RDL
Analyte	mg/l		mg/l	mg/l
Dissolved Solids	U		2.82	10.0

1

Cp

2

Tc

3

Ss

4

Cn

5

Sr

6

Qc

7

Gl

8

Al

9

Sc

L1145752-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1145752-01 10/06/19 14:57 • (DUP) R3458404-3 10/06/19 14:57

	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Analyte	mg/l	mg/l		%		%
Dissolved Solids	1180	1230	1	4.33		5

Laboratory Control Sample (LCS)

(LCS) R3458404-2 10/06/19 14:57

	Spike Amount	LCS Result	LCS Rec.	Rec. Limits	LCS Qualifier
Analyte	mg/l	mg/l	%	%	
Dissolved Solids	8800	8580	97.5	85.0-115	



Method Blank (MB)

(MB) R3457626-1 10/03/19 09:33

	MB Result	MB Qualifier	MB MDL	MB RDL
Analyte	mg/l		mg/l	mg/l
Chloride	U		0.0519	1.00
Sulfate	U		0.0774	5.00

L1145726-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1145726-01 10/03/19 13:15 • (DUP) R3457626-3 10/03/19 13:28

	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Analyte	mg/l	mg/l		%		%
Chloride	618	611	20	1.15		15
Sulfate	1430	1410	20	1.37		15

L1145756-24 Original Sample (OS) • Duplicate (DUP)

(OS) L1145756-24 10/03/19 17:40 • (DUP) R3457626-6 10/03/19 18:19

	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Analyte	mg/l	mg/l		%		%
Chloride	95.4	93.6	1	1.88		15
Sulfate	66.7	65.5	1	1.81		15

Laboratory Control Sample (LCS)

(LCS) R3457626-2 10/03/19 09:46

	Spike Amount	LCS Result	LCS Rec.	Rec. Limits	LCS Qualifier
Analyte	mg/l	mg/l	%	%	
Chloride	40.0	39.2	97.9	80.0-120	
Sulfate	40.0	39.6	98.9	80.0-120	

L1145744-01 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1145744-01 10/03/19 14:07 • (MS) R3457626-4 10/03/19 14:21 • (MSD) R3457626-5 10/03/19 14:34

	Spike Amount	Original Result	MS Result	MSD Result	MS Rec.	MSD Rec.	Dilution	Rec. Limits	MS Qualifier	MSD Qualifier	RPD	RPD Limits
Analyte	mg/l	mg/l	mg/l	mg/l	%	%		%			%	%
Chloride	50.0	129	173	172	88.7	86.9	1	80.0-120	E	E	0.529	15
Sulfate	50.0	44.7	93.3	93.0	97.2	96.7	1	80.0-120			0.262	15

1Cp

2Tc

3Ss

4Cn

5Sr

6Qc

7Gl

8Al

9Sc



L1145760-01 Original Sample (OS) • Matrix Spike (MS)

(OS) L1145760-01 10/03/19 19:50 • (MS) R3457626-7 10/03/19 20:03

Analyte	Spike Amount mg/l	Original Result mg/l	MS Result mg/l	MS Rec. %	Dilution	Rec. Limits %	<u>MS Qualifier</u>
Chloride	50.0	18.6	68.8	100	1	80.0-120	
Sulfate	50.0	18.2	68.3	100	1	80.0-120	

<sup>1</sup>Cp

<sup>2</sup>Tc

<sup>3</sup>Ss

<sup>4</sup>Cn

<sup>5</sup>Sr

<sup>6</sup>Qc

<sup>7</sup>Gl

<sup>8</sup>Al

<sup>9</sup>Sc



Method Blank (MB)

(MB) R3458190-2 10/05/19 11:45

Analyte	MB Result mg/l	MB Qualifier	MB MDL mg/l	MB RDL mg/l
TPH (GC/FID) Low Fraction	0.0778	⬇	0.0314	0.100
(S) a,a,a-Trifluorotoluene(FID)	103			78.0-120

1  
Cp

2  
Tc

3  
Ss

4  
Cn

5  
Sr

6  
Qc

7  
Gl

8  
Al

9  
Sc

Laboratory Control Sample (LCS)

(LCS) R3458190-1 10/05/19 11:00

Analyte	Spike Amount mg/l	LCS Result mg/l	LCS Rec. %	Rec. Limits %	LCS Qualifier
TPH (GC/FID) Low Fraction	5.50	5.44	98.9	72.0-127	
(S) a,a,a-Trifluorotoluene(FID)			105	78.0-120	

L1145883-07 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1145883-07 10/05/19 18:37 • (MS) R3458190-3 10/05/19 18:59 • (MSD) R3458190-4 10/05/19 19:21

Analyte	Spike Amount mg/l	Original Result mg/l	MS Result mg/l	MSD Result mg/l	MS Rec. %	MSD Rec. %	Dilution	Rec. Limits %	MS Qualifier	MSD Qualifier	RPD %	RPD Limits %
TPH (GC/FID) Low Fraction	5.50	ND	5.78	5.71	105	104	1	10.0-160			1.22	22
(S) a,a,a-Trifluorotoluene(FID)					106	106		78.0-120				



Method Blank (MB)

(MB) R3458182-3 10/05/19 04:34

Analyte	MB Result mg/l	MB Qualifier	MB MDL mg/l	MB RDL mg/l
Benzene	U		0.000331	0.00100
Ethylbenzene	U		0.000384	0.00100
Toluene	U		0.000412	0.00100
Xylenes, Total	U		0.00106	0.00300
(S) Toluene-d8	104			80.0-120
(S) 4-Bromofluorobenzene	105			77.0-126
(S) 1,2-Dichloroethane-d4	107			70.0-130

Laboratory Control Sample (LCS) • Laboratory Control Sample Duplicate (LCSD)

(LCS) R3458182-1 10/05/19 03:33 • (LCSD) R3458182-2 10/05/19 03:53

Analyte	Spike Amount mg/l	LCS Result mg/l	LCSD Result mg/l	LCS Rec. %	LCSD Rec. %	Rec. Limits %	LCS Qualifier	LCSD Qualifier	RPD %	RPD Limits %
Benzene	0.0250	0.0248	0.0247	99.2	98.8	70.0-123			0.404	20
Ethylbenzene	0.0250	0.0264	0.0265	106	106	79.0-123			0.378	20
Toluene	0.0250	0.0248	0.0247	99.2	98.8	79.0-120			0.404	20
Xylenes, Total	0.0750	0.0809	0.0808	108	108	79.0-123			0.124	20
(S) Toluene-d8				105	104	80.0-120				
(S) 4-Bromofluorobenzene				104	104	77.0-126				
(S) 1,2-Dichloroethane-d4				113	115	70.0-130				

<sup>1</sup>Cp

<sup>2</sup>Tc

<sup>3</sup>Ss

<sup>4</sup>Cn

<sup>5</sup>Sr

<sup>6</sup>Qc

<sup>7</sup>Gl

<sup>8</sup>Al

<sup>9</sup>Sc



Method Blank (MB)

(MB) R3458266-1 10/06/19 02:42

Analyte	MB Result mg/l	MB Qualifier	MB MDL mg/l	MB RDL mg/l
TPH (GC/FID) High Fraction	U		0.0247	0.100
(S) o-Terphenyl	76.5			31.0-160

<sup>1</sup>Cp

<sup>2</sup>Tc

<sup>3</sup>Ss

<sup>4</sup>Cn

<sup>5</sup>Sr

<sup>6</sup>Qc

<sup>7</sup>Gl

<sup>8</sup>Al

<sup>9</sup>Sc

Laboratory Control Sample (LCS) • Laboratory Control Sample Duplicate (LCSD)

(LCS) R3458266-2 10/06/19 03:02 • (LCSD) R3458266-3 10/06/19 03:23

Analyte	Spike Amount mg/l	LCS Result mg/l	LCSD Result mg/l	LCS Rec. %	LCSD Rec. %	Rec. Limits %	LCS Qualifier	LCSD Qualifier	RPD %	RPD Limits %
TPH (GC/FID) High Fraction	1.50	1.16	1.21	77.3	80.7	50.0-150			4.22	20
(S) o-Terphenyl				92.0	93.0	31.0-160				

L1145696-01 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1145696-01 10/06/19 11:18 • (MS) R3458266-4 10/06/19 11:38 • (MSD) R3458266-5 10/06/19 11:58

Analyte	Spike Amount mg/l	Original Result mg/l	MS Result mg/l	MSD Result mg/l	MS Rec. %	MSD Rec. %	Dilution	Rec. Limits %	MS Qualifier	MSD Qualifier	RPD %	RPD Limits %
TPH (GC/FID) High Fraction	1.43	ND	1.46	1.52	102	106	1	50.0-150			4.03	20
(S) o-Terphenyl					87.9	86.3		31.0-160				



## Guide to Reading and Understanding Your Laboratory Report

The information below is designed to better explain the various terms used in your report of analytical results from the Laboratory. This is not intended as a comprehensive explanation, and if you have additional questions please contact your project representative.

Results Disclaimer - Information that may be provided by the customer, and contained within this report, include Permit Limits, Project Name, Sample ID, Sample Matrix, Sample Preservation, Field Blanks, Field Spikes, Field Duplicates, On-Site Data, Sampling Collection Dates/Times, and Sampling Location. Results relate to the accuracy of this information provided, and as the samples are received.

### Abbreviations and Definitions

MDL	Method Detection Limit.
ND	Not detected at the Reporting Limit (or MDL where applicable).
RDL	Reported Detection Limit.
Rec.	Recovery.
RPD	Relative Percent Difference.
SDG	Sample Delivery Group.
(S)	Surrogate (Surrogate Standard) - Analytes added to every blank, sample, Laboratory Control Sample/Duplicate and Matrix Spike/Duplicate; used to evaluate analytical efficiency by measuring recovery. Surrogates are not expected to be detected in all environmental media.
U	Not detected at the Reporting Limit (or MDL where applicable).
Analyte	The name of the particular compound or analysis performed. Some Analyses and Methods will have multiple analytes reported.
Dilution	If the sample matrix contains an interfering material, the sample preparation volume or weight values differ from the standard, or if concentrations of analytes in the sample are higher than the highest limit of concentration that the laboratory can accurately report, the sample may be diluted for analysis. If a value different than 1 is used in this field, the result reported has already been corrected for this factor.
Limits	These are the target % recovery ranges or % difference value that the laboratory has historically determined as normal for the method and analyte being reported. Successful QC Sample analysis will target all analytes recovered or duplicated within these ranges.
Original Sample	The non-spiked sample in the prep batch used to determine the Relative Percent Difference (RPD) from a quality control sample. The Original Sample may not be included within the reported SDG.
Qualifier	This column provides a letter and/or number designation that corresponds to additional information concerning the result reported. If a Qualifier is present, a definition per Qualifier is provided within the Glossary and Definitions page and potentially a discussion of possible implications of the Qualifier in the Case Narrative if applicable.
Result	The actual analytical final result (corrected for any sample specific characteristics) reported for your sample. If there was no measurable result returned for a specific analyte, the result in this column may state "ND" (Not Detected) or "BDL" (Below Detectable Levels). The information in the results column should always be accompanied by either an MDL (Method Detection Limit) or RDL (Reporting Detection Limit) that defines the lowest value that the laboratory could detect or report for this analyte.
Uncertainty (Radiochemistry)	Confidence level of 2 sigma.
Case Narrative (Cn)	A brief discussion about the included sample results, including a discussion of any non-conformances to protocol observed either at sample receipt by the laboratory from the field or during the analytical process. If present, there will be a section in the Case Narrative to discuss the meaning of any data qualifiers used in the report.
Quality Control Summary (Qc)	This section of the report includes the results of the laboratory quality control analyses required by procedure or analytical methods to assist in evaluating the validity of the results reported for your samples. These analyses are not being performed on your samples typically, but on laboratory generated material.
Sample Chain of Custody (Sc)	This is the document created in the field when your samples were initially collected. This is used to verify the time and date of collection, the person collecting the samples, and the analyses that the laboratory is requested to perform. This chain of custody also documents all persons (excluding commercial shippers) that have had control or possession of the samples from the time of collection until delivery to the laboratory for analysis.
Sample Results (Sr)	This section of your report will provide the results of all testing performed on your samples. These results are provided by sample ID and are separated by the analyses performed on each sample. The header line of each analysis section for each sample will provide the name and method number for the analysis reported.
Sample Summary (Ss)	This section of the Analytical Report defines the specific analyses performed for each sample ID, including the dates and times of preparation and/or analysis.

Qualifier	Description
E	The analyte concentration exceeds the upper limit of the calibration range of the instrument established by the initial calibration (ICAL).
J	The identification of the analyte is acceptable; the reported value is an estimate.

1	Cp
2	Tc
3	Ss
4	Cn
5	Sr
6	Qc
7	Gl
8	Al
9	Sc



Pace National is the only environmental laboratory accredited/certified to support your work nationwide from one location. One phone call, one point of contact, one laboratory. No other lab is as accessible or prepared to handle your needs throughout the country. Our capacity and capability from our single location laboratory is comparable to the collective totals of the network laboratories in our industry. The most significant benefit to our one location design is the design of our laboratory campus. The model is conducive to accelerated productivity, decreasing turn-around time, and preventing cross contamination, thus protecting sample integrity. Our focus on premium quality and prompt service allows us to be YOUR LAB OF CHOICE.

\* Not all certifications held by the laboratory are applicable to the results reported in the attached report.

\* Accreditation is only applicable to the test methods specified on each scope of accreditation held by Pace National.

## State Accreditations

Alabama	40660	Nebraska	NE-OS-15-05
Alaska	17-026	Nevada	TN-03-2002-34
Arizona	AZ0612	New Hampshire	2975
Arkansas	88-0469	New Jersey–NELAP	TN002
California	2932	New Mexico <sup>1</sup>	n/a
Colorado	TN00003	New York	11742
Connecticut	PH-0197	North Carolina	Env375
Florida	E87487	North Carolina <sup>1</sup>	DW21704
Georgia	NELAP	North Carolina <sup>3</sup>	41
Georgia <sup>1</sup>	923	North Dakota	R-140
Idaho	TN00003	Ohio–VAP	CL0069
Illinois	200008	Oklahoma	9915
Indiana	C-TN-01	Oregon	TN200002
Iowa	364	Pennsylvania	68-02979
Kansas	E-10277	Rhode Island	LA000356
Kentucky <sup>1 6</sup>	90010	South Carolina	84004
Kentucky <sup>2</sup>	16	South Dakota	n/a
Louisiana	AI30792	Tennessee <sup>1 4</sup>	2006
Louisiana <sup>1</sup>	LA180010	Texas	T104704245-18-15
Maine	TN0002	Texas <sup>5</sup>	LAB0152
Maryland	324	Utah	TN00003
Massachusetts	M-TN003	Vermont	VT2006
Michigan	9958	Virginia	460132
Minnesota	047-999-395	Washington	C847
Mississippi	TN00003	West Virginia	233
Missouri	340	Wisconsin	9980939910
Montana	CERT0086	Wyoming	A2LA

## Third Party Federal Accreditations

A2LA – ISO 17025	1461.01	AIHA-LAP, LLC EMLAP	100789
A2LA – ISO 17025 <sup>5</sup>	1461.02	DOD	1461.01
Canada	1461.01	USDA	P330-15-00234
EPA–Crypto	TN00003		

<sup>1</sup> Drinking Water <sup>2</sup> Underground Storage Tanks <sup>3</sup> Aquatic Toxicity <sup>4</sup> Chemical/Microbiological <sup>5</sup> Mold <sup>6</sup> Wastewater n/a Accreditation not applicable

## Our Locations

Pace National has sixty-four client support centers that provide sample pickup and/or the delivery of sampling supplies. If you would like assistance from one of our support offices, please contact our main office. Pace National performs all testing at our central laboratory.



