

Session Report

3/18/2020

Information Panel


Name S059_BGN010010_18032020_132711
Start Time 3/16/2020 1:16:03 PM
Stop Time 3/18/2020 11:57:56 AM
Device Name BGN010010
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	49.2 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	FAST	Bandwidth	1	OFF
Leq	2	68.2 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Logged Data Chart

S059_BGN010010_18032020_132711: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
3/16/2020 1:15:17 PM	Calibration	114.0			