

Session Report

5/21/2020

General Information

Name S136_BGN010009_21052020_085124

Comments

Start Time 5/14/2020 9:26:26 AM

Stop Time 5/14/2020 9:45:29 AM

Run Time 00:19:03

Model Type SoundPro DL

Serial Number BGN010009

Device Firmware Rev R.13F

Company Name

Description

Location

User Name

Summary Data

Description	Meter	Value	Description	Meter	Value
Dose	1	0 %	Pdose (8:00)	1	0.1 %
Lavg	1	--	Lpk	1	83.4 dB
Leq	1	52.8 dB	TWA	1	38.8 dB
UL Time	1	00:00:00	SEL	1	83.4 dB
ProjectedTWA (8:00)	1	52.8 dB	Mntime	1	5/14/2020 9:27:44 AM
Mxtime	1	5/14/2020 9:30:01 AM	PKtime	1	5/14/2020 9:26:47 AM
Weighting	1	--	Range Ceiling	1	--
Criterion Level	1	--	ULL	1	--
Dynamic Range	1	--	Exchange Rate	1	--
Response	1	--	Int Threshold	1	--
Alarm Level 1	1	--	AlarmLevel2	1	--
Dosimeter Name	1	--			
Dose	2	0 %	Pdose (8:00)	2	0.1 %
Lavg	2	--	Lpk	2	83.4 dB
Leq	2	52.8 dB	TWA	2	38.7 dB
UL Time	2	00:00:00	SEL	2	83.3 dB

ProjectedTWA (8:00)	2	52.8 dB	Mntime	2	5/14/2020 9:27:41 AM
Mxtime	2	5/14/2020 9:26:47 AM	PKtime	2	5/14/2020 9:26:47 AM
Weighting	2	C	Range Ceiling	2	--
Criterion Level	2	85 dB	ULL	2	115 dB
Dynamic Range	2	--	Exchange Rate	2	3 dB
Response	2	SLOW	Integrating Threshold	2	100 dB
Alarm Level 1	2	--	AlarmLevel2	2	--
Dosimeter Name	2	--			

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
5/14/2020 9:23:15 AM	Calibration	114.0			
5/14/2020 9:47:04 AM	Verification	114.1			

Logged Data Chart

S136_BGN010009_21052020_085124: Logged Data Chart - Read Only

