

Session Report

8/25/2020

General Information

Name S137_BGN010009_25082020_084454

Comments

Start Time 7/23/2020 10:19:22 AM

Stop Time 7/23/2020 10:42:24 AM

Run Time 00:23:02

Model Type SoundPro DL

Serial Number BGN010009

Device Firmware Rev R.13F

Company Name

Description

Location

User Name

Summary Data

Description	Meter	Value	Description	Meter	Value
Dose	1	0 %	Pdose (8:00)	1	0.3 %
Lavg	1	--	Lpk	1	74.8 dB
Leq	1	59.6 dB	TWA	1	46.4 dB
UL Time	1	00:00:00	SEL	1	91 dB
ProjectedTWA (8:00)	1	59.6 dB	Mntime	1	7/23/2020 10:40:40 AM
Mxtime	1	7/23/2020 10:19:31 AM	PKtime	1	7/23/2020 10:19:31 AM
Weighting	1	--	Range Ceiling	1	--
Criterion Level	1	--	ULL	1	--
Dynamic Range	1	--	Exchange Rate	1	--
Response	1	--	Int Threshold	1	--
Alarm Level 1	1	--	AlarmLevel2	1	--
Dosimeter Name	1	--			
Dose	2	0.1 %	Pdose (8:00)	2	2.2 %
Lavg	2	--	Lpk	2	85.1 dB
Leq	2	68.4 dB	TWA	2	55.2 dB
UL Time	2	00:00:00	SEL	2	99.8 dB

ProjectedTWA (8:00)	2	68.4 dB	Mntime	2	7/23/2020 10:39:52 AM
Mxtime	2	7/23/2020 10:19:30 AM	PKtime	2	7/23/2020 10:19:28 AM
Weighting	2	C	Range Ceiling	2	--
Criterion Level	2	85 dB	ULL	2	115 dB
Dynamic Range	2	--	Exchange Rate	2	3 dB
Response	2	SLOW	Integrating Threshold	2	100 dB
Alarm Level 1	2	--	AlarmLevel2	2	--
Dosimeter Name	2	--			

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
7/23/2020 10:17:48 AM	Calibration	114.0			
7/23/2020 10:43:30 AM	Verification	113.7			

Logged Data Chart

S137_BGN010009_25082020_084454: Logged Data Chart - Read Only

