

# Session Report

8/25/2020

## General Information

Name S138\_BGN010009\_25082020\_084455

Comments

Start Time 8/20/2020 8:38:13 AM

Stop Time 8/20/2020 9:00:07 AM

Run Time 00:21:54

Model Type SoundPro DL

Serial Number BGN010009

Device Firmware Rev R.13F

Company Name

Description

Location

User Name

## Summary Data

Description	Meter	Value	Description	Meter	Value
Dose	1	0 %	Pdose (8:00)	1	0.2 %
Lavg	1	--	Lpk	1	82.7 dB
Leq	1	56.9 dB	TWA	1	43.5 dB
UL Time	1	00:00:00	SEL	1	88.1 dB
ProjectedTWA (8:00)	1	56.9 dB	Mntime	1	8/20/2020 8:51:01 AM
Mxtime	1	8/20/2020 8:50:09 AM	PKtime	1	8/20/2020 8:38:48 AM
Weighting	1	--	Range Ceiling	1	--
Criterion Level	1	--	ULL	1	--
Dynamic Range	1	--	Exchange Rate	1	--
Response	1	--	Int Threshold	1	--
Alarm Level 1	1	--	AlarmLevel2	1	--
Dosimeter Name	1	--			
Dose	2	0 %	Pdose (8:00)	2	1 %
Lavg	2	--	Lpk	2	80.8 dB
Leq	2	65 dB	TWA	2	51.6 dB
UL Time	2	00:00:00	SEL	2	96.2 dB

ProjectedTWA (8:00)	2	65 dB	Mntime	2	8/20/2020 8:51:01 AM
Mxtime	2	8/20/2020 9:00:04 AM	PKtime	2	8/20/2020 8:38:48 AM
Weighting	2	C	Range Ceiling	2	--
Criterion Level	2	85 dB	ULL	2	115 dB
Dynamic Range	2	--	Exchange Rate	2	3 dB
Response	2	SLOW	Integrating Threshold	2	100 dB
Alarm Level 1	2	--	AlarmLevel2	2	--
Dosimeter Name	2	--			

## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
8/20/2020 8:35:57 AM	Calibration	114.0			
8/20/2020 9:02:16 AM	Verification	113.8			

## Logged Data Chart

S138\_BGN010009\_25082020\_084455: Logged Data Chart - Read Only

