

Session Report


2/22/2021

Information Panel

Name S046_BGN010014_22022021_111827
Start Time 2/19/2021 10:19:03 AM
Stop Time 2/22/2021 10:44:52 AM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments See weather data attachment for exclusion details.

Logged Data Chart

S046_BGN010014_22022021_111827: Logged Data Chart



Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	58.1 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	68.6 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
2/19/2021 10:17:15 AM	Calibration	114.0			
2/22/2021 10:45:19 AM	Verification	114.0			