

# Session Report


4/19/2021

## Information Panel

Name S049\_BGN010014\_19042021\_105427  
Start Time 4/9/2021 9:11:21 AM  
Stop Time 4/16/2021 8:53:17 AM  
Device Name BGN010014  
Model Type SoundPro DL  
Device Firmware Rev R.13F  
Comments

## Logged Data Chart

S049\_BGN010014\_19042021\_105427: Logged Data Chart



## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
4/9/2021 9:10:55 AM	Calibration	114.0			
4/16/2021 8:54:12 AM	Verification	114.3			

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	48.9 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	68.1 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			