

Session Report

8/6/2021

Information Panel


Name S055_BGN010014_06082021_132637
Start Time 8/6/2021 11:22:41 AM
Stop Time 8/6/2021 11:55:17 AM
Device Name BGN010014
Model Type SoundPro DL
Device Firmware Rev R.13F
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	56.9 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	SLOW	Bandwidth	1	OFF
Leq	2	75.5 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			

Logged Data Chart

S055_BGN010014_06082021_132637: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
8/6/2021 11:22:09 AM	Calibration	114.0			
8/6/2021 11:55:42 AM	Verification	113.7			