



Tuesday, May 31, 2022

Max Trehus  
Great Western Operating Company, LLC  
4093 Specialty Place, Unit B  
Longmont, CO 80504

Re: ALS Workorder: 2205376  
Project Name: Postle IC 09-299HNBH  
Project Number:

Dear Mr. Trehus:

Two water samples were received from Great Western Operating Company, LLC, on 5/18/2022. The samples were scheduled for the following analyses:

- Dissolved Gasses
- GC/MS Volatiles
- Inorganics
- Metals
- Total Extractable Petroleum Hydrocarbons (Diesel)
- Total Volatile Petroleum Hydrocarbons (Gasoline)

The results for these analyses are contained in the enclosed reports.

The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, ALS certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed. Should this laboratory report need to be reproduced, it should be reproduced in full unless written approval has been obtained from ALS Environmental.

Thank you for your confidence in ALS Environmental. Should you have any questions, please call.

Sincerely,

For  
ALS Environmental  
Katie M. OBrien  
Project Manager

Accreditations: ALS Environmental – Fort Collins is accredited by the following accreditation bodies for various testing scopes in accordance with requirements of each accreditation body. All testing is performed under the laboratory management system, which is maintained to meet these requirement and regulations. Please contact the laboratory or accreditation body for the current scope testing parameters.

ALS Environmental – Fort Collins	
Accreditation Body	License or Certification Number
Arizona	AZ0828
California (CA)	2926
Colorado (CO)	CO01099
Florida (FL)	E87914
Idaho (ID)	CO01099
Kansas (KS)	E-10381
Kentucky (KY)	90137
Oklahoma	1301
Maryland (MD)	285
Missouri (MO)	175
Nebraska(NE)	NE-OS-24-13
Nevada (NV)	CO010992018-1
New York (NY)	12036
North Dakota (ND)	R-057
Oklahoma (OK)	1301
Pennsylvania (PA)	68-03116
Tennessee (TN)	TN02976
Texas (TX)	T104704241
Utah (UT)	CO01099
Washington (WA)	C1280
Virginia	460305

40 CFR Part 136: All analyses for Clean Water Act samples are analyzed using the 40 CFR Part 136 specified method and include all the QC requirements.



## 2205376

### GC/MS Volatiles:

The sample was analyzed using GC/MS following the current revision of SOP 525 based on SW-846 Method 8260C.

All surrogate recoveries were within acceptance criteria with the following exception:

Surrogate	Sample	Direction
Dibromofluoromethane	-1	Low

The low surrogate recovery is likely due to the high pH of the sample. No further action was taken.

All remaining acceptance criteria were met.

### Dissolved Gasses:

The sample was prepared and analyzed according to method RSK-175 procedures and the current revision of SOP 449.

All acceptance criteria were met.

### GRO:

The sample was analyzed following the current revision of SOP 425 generally based on SW-846 Methods 8000C and 8015D. TVPH is a multicomponent mixture and is quantitated by summing the entire carbon range, rather than individual peaks. The carbon range integrated in this test extends from C6 to C10.

All acceptance criteria were met.

### DRO:

The sample was analyzed following the current revision of SOP 406 generally based on SW-846 Methods 8000C and 8015D. TEPH is a multicomponent mixture and is quantitated by summing the entire carbon range, rather than individual peaks. The carbon range integrated in this test extends from C10 to C28.

All acceptance criteria were met.



**Metals:**

The samples were analyzed following Methods for the Determination of Metals in Environmental Samples – Supplement 1 procedures. Analysis by Trace ICP followed method 200.7 and the current revision of SOP 834.

Sample 2205376-2 was to be analyzed for dissolved metals. The sample was filtered through a 0.45 micron filter and preserved with nitric acid to a pH less than two prior to analysis.

All acceptance criteria were met.

**Inorganics:**

The sample was analyzed following EMSL and Standard Method procedures for the current revisions of the following SOPs and methods:

<u>Analyte</u>	<u>Method</u>	<u>SOP #</u>
Alkalinity	SM2320B	1106
Bicarbonate	SM2320B	1106
Carbonate	SM2320B	1106
TDS	SM2540C	1101
Chloride	300.0 Revision 2.1	1113
Sulfate	300.0 Revision 2.1	1113

All acceptance criteria were met.

# ALS -- Fort Collins

## Sample Number(s) Cross-Reference Table

---

**OrderNum:** 2205376

**Client Name:** PDC Energy

**Client Project Name:** Postle IC 09-299HNBH

**Client Project Number:**

**Client PO Number:**

---

Client Sample Number	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
09-299HNA	2205376-1		WATER	18-May-22	12:23
09-299HNB	2205376-2		WATER	18-May-22	12:23



**ALS Environmental**

225 Commerce Drive, Fort Collins, Colorado 80524  
 TF: (800) 443-1511 PH: (970) 490-1511 FX: (970) 490-1522

**Chain-of-Custody**

Form 202a

WORKORDER # **2205376**

PROJECT NAME	Postle IL 09-299HN BH	SAMPLER	Max Trehus	DATE		PAGE		of
PROJECT No.		SITE ID		TURNAROUND		DISPOSAL		By Lab or Return to Client
COMPANY NAME	DDC	EDD FORMAT						
SEND REPORT TO	Max Trehus	PURCHASE ORDER						
ADDRESS	Jenifer Hakkarinen	BILL TO COMPANY						
CITY/STATE/ZIP	Jessica Johansen	INVOICE ATTN TO						
PHONE	max.trehus@pdce.com	ADDRESS						
FAX	Jenifer.Hakkarinen@pdce.com	CITY/STATE/ZIP						
E-MAIL	Jessica.Johansen@pdce.com	PHONE						
		FAX						
		E-MAIL						

Lab ID	Field ID	Matrix	Sample Date	Sample Time	# Bottles	Pres.	QC	TURNAROUND	DATE	By Lab or Return to Client
	09-299HN A	W	5/18	12:23	3	-		Disolved Gasses		
	09-299HNA				3	1		DS metal		
	09-299HNA				3	1		Amions, Hk, TDS		
	09-299HN A				3	1		gco		
	09-299HNA				1	-		PRO		
2	09-299HN B				1	-		STEX		
1	09-299HN A				1	2				

\*Time Zone (Circle): EST CST MST PST Matrix: O = oil S = soil NS = non-soil solid W = water L = liquid E = extract F = filter

For metals or anions, please detail analytes below.

Comments: Facility ID: 4533A7 S/C

RELINQUISHED BY	Signature	PRINTED NAME	Max Trehus	DATE	5/18	TIME	14:48
RECEIVED BY	Signature	PRINTED NAME	Cherie Thomas	DATE	5/18	TIME	1500
RELINQUISHED BY							
RECEIVED BY							
RELINQUISHED BY							
RECEIVED BY							

Preservative Key: 1-HCl 2-HNO3 3-H2SO4 4-NaOH 5-NaHSO4 7-Other 8-4 degrees C 9-5035



**ALS Environmental - Fort Collins**  
**CONDITION OF SAMPLE UPON RECEIPT FORM**

Client: PDC

Workorder No: 2205376

Project Manager: \_\_\_\_\_

Initials: CXT

Date: 05/18/2022

		N/A	YES	NO
1.	Are airbills / shipping documents present and/or removable?	X		
	Tracking number: _____			
2.	Are custody seals on shipping containers intact?	X		
3.	Are custody seals on sample containers intact?	X		
4.	Is there a COC (chain-of-custody) present?		X	
5.	Is the COC in agreement with samples received? (IDs, dates, times, # of samples, # of containers, matrix, requested analyses, etc.)		X	
6.	Are short-hold samples present?			X
7.	Are all samples within holding times for the requested analyses?		X	
8.	Were all sample containers received intact? (not broken or leaking)		X	
9.	Is there sufficient sample for the requested analyses?		X	
10.	Are samples in proper containers for requested analyses? (form 250, <i>Sample Handling Guidelines</i> )		X	
11.	Are all aqueous samples preserved correctly, if required? (excluding volatiles)		X	
12.	Are all samples requiring no headspace (VOC, GRO, RSK/MEE, radon) free of bubbles > 6 mm (1/4 inch) diameter? (i.e. size of green pea)			X
13.	Were the samples shipped on ice?		X	
14.	Were cooler temperatures measured at 0.1-6.0°C?	IR gun used*: #6	RAD ONLY	X
Cooler #: <u>1</u> Temperature (°C): <u>5.5</u> # of custody seals on cooler: <u>0</u> External µR/hr reading: <u>NA</u> Background µR/hr reading: <u>11</u> Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <u>NA</u> (If no, see Form 008.)				

\* Please provide details here for NO responses to boxes above - for 2 thru 5 & 7 thru 12, notify PM & continue w/ login.

**Sample 1 bottles 1-12 received with headspace**

---

---

---

---

---

---

---

---

Were unpreserved bottles pH checked? NA

All client bottle ID's vs ALS lab ID's double-checked by CT

If applicable, was the client contacted? YES / NO / NA Contact: \_\_\_\_\_

Date/Time: \_\_\_\_\_

Project Manager Signature / Date: 

5/19/22

**ALS -- Fort Collins**

**SAMPLE SUMMARY REPORT**

**Client:** PDC Energy  
**Project:** Postle IC 09-299HNBH  
**Sample ID:** 09-299HNA  
**Legal Location:**  
**Collection Date:** 5/18/2022 12:23

**Date:** 31-May-22  
**Work Order:** 2205376  
**Lab ID:** 2205376-1  
**Matrix:** WATER  
**Percent Moisture:**

Analyses	Result	Qual	Report Limit	Units	Dilution Factor	Date Analyzed
<b>Alkalinity as Calcium Carbonate</b>			<b>SM2320B</b>			Prep Date: <b>5/26/2022</b> PrepBy: <b>KRL</b>
BICARBONATE AS CaCO3	ND		20	MG/L	1	5/26/2022
<b>CARBONATE AS CaCO3</b>	<b>1500</b>		<b>20</b>	<b>MG/L</b>	1	5/26/2022
<b>TOTAL ALKALINITY AS CaCO3</b>	<b>11000</b>		<b>20</b>	<b>MG/L</b>	1	5/26/2022
<b>Diesel Range Organics</b>			<b>SW8015M</b>			Prep Date: <b>5/20/2022</b> PrepBy: <b>JRS</b>
Diesel Range Organics	73		1	MG/L	1	5/23/2022 23:49
Surr: O-TERPHENYL	99		69-120	%REC	1	5/23/2022 23:49
<b>Dissolved Gasses</b>			<b>RSK175</b>			Prep Date: <b>5/20/2022</b> PrepBy: <b>JRS</b>
<b>METHANE</b>	<b>1100</b>		<b>1</b>	<b>UG/L</b>	1	5/23/2022 16:20
<b>ETHANE</b>	<b>540</b>		<b>2</b>	<b>UG/L</b>	1	5/23/2022 16:20
<b>PROPANE</b>	<b>600</b>		<b>1</b>	<b>UG/L</b>	1	5/23/2022 16:20
<b>Gasoline Range Organics</b>			<b>SW8015</b>			Prep Date: <b>5/20/2022</b> PrepBy: <b>JRS</b>
<b>GASOLINE RANGE ORGANICS</b>	<b>9.3</b>		<b>1</b>	<b>MG/L</b>	10	5/25/2022 21:15
Surr: 2,3,4-TRIFLUOROTOLUENE	80		80-120	%REC	10	5/25/2022 21:15
<b>GC/MS Volatiles</b>			<b>SW8260_25</b>			Prep Date: <b>5/26/2022</b> PrepBy: <b>TWK</b>
<b>BENZENE</b>	<b>58</b>		<b>50</b>	<b>UG/L</b>	50	5/27/2022 01:47
<b>TOLUENE</b>	<b>120</b>		<b>50</b>	<b>UG/L</b>	50	5/27/2022 01:47
<b>ETHYLBENZENE</b>	<b>600</b>		<b>50</b>	<b>UG/L</b>	50	5/27/2022 01:47
<b>M+P-XYLENE</b>	<b>2300</b>		<b>50</b>	<b>UG/L</b>	50	5/27/2022 01:47
<b>O-XYLENE</b>	<b>600</b>		<b>50</b>	<b>UG/L</b>	50	5/27/2022 01:47
<b>TOTAL XYLENES</b>	<b>2900</b>		<b>1</b>	<b>UG/L</b>	1	5/27/2022 01:47
Surr: 4-BROMOFLUOROBENZENE	100		80-120	%REC	50	5/27/2022 01:47
Surr: DIBROMOFLUOROMETHANE	76	*	80-120	%REC	50	5/27/2022 01:47
Surr: TOLUENE-D8	95		80-120	%REC	50	5/27/2022 01:47
<b>Ion Chromatography</b>			<b>EPA300.0</b>			Prep Date: <b>5/26/2022</b> PrepBy: <b>AOW</b>
<b>CHLORIDE</b>	<b>1600</b>		<b>100</b>	<b>MG/L</b>	500	5/26/2022 15:24
<b>SULFATE</b>	<b>590</b>		<b>500</b>	<b>MG/L</b>	500	5/26/2022 15:24
<b>Total Recoverable Metals by 200.7</b>			<b>EPA200.7</b>			Prep Date: <b>5/23/2022</b> PrepBy: <b>ETC</b>
<b>CALCIUM</b>	<b>98</b>		<b>10</b>	<b>MG/L</b>	10	5/25/2022 09:28
<b>POTASSIUM</b>	<b>6500</b>		<b>100</b>	<b>MG/L</b>	100	5/25/2022 10:23
<b>MAGNESIUM</b>	<b>ND</b>		<b>10</b>	<b>MG/L</b>	10	5/25/2022 09:28
<b>SODIUM</b>	<b>1900</b>		<b>10</b>	<b>MG/L</b>	10	5/25/2022 09:28
<b>Total Dissolved Solids</b>			<b>SM2540C</b>			Prep Date: <b>5/23/2022</b> PrepBy: <b>AOW</b>
<b>TOTAL DISSOLVED SOLIDS</b>	<b>18000</b>		<b>1000</b>	<b>MG/L</b>	1	5/25/2022

**Client:** PDC Energy  
**Project:** Postle IC 09-299HNBH  
**Sample ID:** 09-299HNB  
**Legal Location:**  
**Collection Date:** 5/18/2022 12:23

**Date:** 31-May-22  
**Work Order:** 2205376  
**Lab ID:** 2205376-2  
**Matrix:** WATER  
**Percent Moisture:**

Analyses	Result	Qual	Report Limit	Units	Dilution Factor	Date Analyzed
<b>Dissolved Metals by 200.7</b>			<b>EPA200.7</b>		Prep Date: <b>5/23/2022</b>	PrepBy: <b>ETC</b>
<b>CALCIUM</b>	<b>94</b>		<b>10</b>	<b>MG/L</b>	10	5/25/2022 09:29
<b>POTASSIUM</b>	<b>6600</b>		<b>100</b>	<b>MG/L</b>	100	5/25/2022 10:24
<b>MAGNESIUM</b>	<b>ND</b>		<b>10</b>	<b>MG/L</b>	10	5/25/2022 09:29
<b>SODIUM</b>	<b>1900</b>		<b>10</b>	<b>MG/L</b>	10	5/25/2022 09:29

**Client:** PDC Energy  
**Project:** Postle IC 09-299HNBH  
**Sample ID:** 09-299HNB  
**Legal Location:**  
**Collection Date:** 5/18/2022 12:23

**Date:** 31-May-22  
**Work Order:** 2205376  
**Lab ID:** 2205376-2  
**Matrix:** WATER  
**Percent Moisture:**

Analyses	Result	Qual	Report Limit	Units	Dilution Factor	Date Analyzed
----------	--------	------	--------------	-------	-----------------	---------------

**Explanation of Qualifiers**

**Radiochemistry:**

- "Report Limit" is the MDC
- U or ND - Result is less than the sample specific MDC.
- Y1 - Chemical Yield is in control at 100-110%. Quantitative yield is assumed.
- Y2 - Chemical Yield outside default limits.
- W - DER is greater than Warning Limit of 1.42
- \* - Aliquot Basis is 'As Received' while the Report Basis is 'Dry Weight'.
- # - Aliquot Basis is 'Dry Weight' while the Report Basis is 'As Received'.
- G - Sample density differs by more than 15% of LCS density.
- D - DER is greater than Control Limit
- M - Requested MDC not met.
- M3 - The requested MDC was not met, but the reported activity is greater than the reported MDC.
- L - LCS Recovery below lower control limit.
- H - LCS Recovery above upper control limit.
- P - LCS, Matrix Spike Recovery within control limits.
- N - Matrix Spike Recovery outside control limits
- NC - Not Calculated for duplicate results less than 5 times MDC
- B - Analyte concentration greater than MDC.
- B3 - Analyte concentration greater than MDC but less than Requested MDC.

**Inorganics:**

- B - Result is less than the requested reporting limit but greater than the instrument method detection limit (MDL).
- U or ND - Indicates that the compound was analyzed for but not detected.
- E - The reported value is estimated because of the presence of interference. An explanatory note may be included in the narrative.
- M - Duplicate injection precision was not met.
- N - Spiked sample recovery not within control limits. A post spike is analyzed for all ICP analyses when the matrix spike and or spike duplicate fail and the native sample concentration is less than four times the spike added concentration.
- Z - Spiked recovery not within control limits. An explanatory note may be included in the narrative.
- \* - Duplicate analysis (relative percent difference) not within control limits.
- S - SAR value is estimated as one or more analytes used in the calculation were not detected above the detection limit.

**Organics:**

- U or ND - Indicates that the compound was analyzed for but not detected.
- B - Analyte is detected in the associated method blank as well as in the sample. It indicates probable blank contamination and warns the data user.
- E - Analyte concentration exceeds the upper level of the calibration range.
- J - Estimated value. The result is less than the reporting limit but greater than the instrument method detection limit (MDL).
- A - A tentatively identified compound is a suspected aldol-condensation product.
- X - The analyte was diluted below an accurate quantitation level.
- \* - The spike recovery is equal to or outside the control criteria used.
- + - The relative percent difference (RPD) equals or exceeds the control criteria.
- G - A pattern resembling gasoline was detected in this sample.
- D - A pattern resembling diesel was detected in this sample.
- M - A pattern resembling motor oil was detected in this sample.
- C - A pattern resembling crude oil was detected in this sample.
- 4 - A pattern resembling JP-4 was detected in this sample.
- 5 - A pattern resembling JP-5 was detected in this sample.
- H - Indicates that the fuel pattern was in the heavier end of the retention time window for the analyte of interest.
- L - Indicates that the fuel pattern was in the lighter end of the retention time window for the analyte of interest.
- Z - This flag indicates that a significant fraction of the reported result did not resemble the patterns of any of the following petroleum hydrocarbon products:
  - gasoline
  - JP-8
  - diesel
  - mineral spirits
  - motor oil
  - Stoddard solvent
  - bunker C

ALS -- Fort Collins

Date: 5/31/2022 8:21:2

Client: PDC Energy  
 Work Order: 2205376  
 Project: Postle IC 09-299HNBH

QC BATCH REPORT

Batch ID: **HC220520-61-1** Instrument ID: **FUELS-1** Method: **SW8015**

LCS		Sample ID: <b>HC220520-61</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/23/2022 14:38</b>				
Client ID:		Run ID: <b>HC220526-61A</b>			Prep Date: <b>5/20/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
GASOLINE RANGE ORGANICS	0.458	0.1	0.5		92	80-120				20	
Surr: 2,3,4-TRIFLUOROTOLUENE	0.0983		0.1		98	80-120					

LCSD		Sample ID: <b>HC220520-61</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/23/2022 17:47</b>				
Client ID:		Run ID: <b>HC220526-61A</b>			Prep Date: <b>5/20/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
GASOLINE RANGE ORGANICS	0.464	0.1	0.5		93	80-120		0.458	1	20	
Surr: 2,3,4-TRIFLUOROTOLUENE	0.107		0.1		107	80-120			8		

MB		Sample ID: <b>HC220520-61</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/23/2022 14:54</b>				
Client ID:		Run ID: <b>HC220526-61A</b>			Prep Date: <b>5/20/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
GASOLINE RANGE ORGANICS	ND	0.1									
Surr: 2,3,4-TRIFLUOROTOLUENE	0.106				106	80-120					

The following samples were analyzed in this batch:

Client: PDC Energy  
 Work Order: 2205376  
 Project: Postle IC 09-299HNBH

# QC BATCH REPORT

Batch ID: **HC220520-81-1** Instrument ID: **FUELS-1** Method: **SW8015M**

LCS		Sample ID: <b>HC220520-81</b>					Units: <b>MG/L</b>	Analysis Date: <b>5/23/2022 16:39</b>				
Client ID:		Run ID: <b>HC220523-81A</b>				Prep Date: <b>5/20/2022</b>			DF: <b>1</b>			
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual	
Diesel Range Organics	9.07	1.07	8.33		109	53-120				20		
Surr: O-TERPHENYL	1.65		1.67		99	69-120						

LCSD		Sample ID: <b>HC220520-81</b>					Units: <b>MG/L</b>	Analysis Date: <b>5/23/2022 17:01</b>				
Client ID:		Run ID: <b>HC220523-81A</b>				Prep Date: <b>5/20/2022</b>			DF: <b>1</b>			
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual	
Diesel Range Organics	9	1.07	8.33		108	53-120		9.07	1	20		
Surr: O-TERPHENYL	1.63		1.67		98	69-120			1			

MB		Sample ID: <b>HC220520-81</b>					Units: <b>MG/L</b>	Analysis Date: <b>5/23/2022 15:57</b>				
Client ID:		Run ID: <b>HC220523-81A</b>				Prep Date: <b>5/20/2022</b>			DF: <b>1</b>			
Analyte	Result	ReportLimit										Qual
Diesel Range Organics	ND	1.1										
Surr: O-TERPHENYL	1.63				98	69-120						

The following samples were analyzed in this batch:

Client: PDC Energy  
 Work Order: 2205376  
 Project: Postle IC 09-299HNBH

# QC BATCH REPORT

Batch ID: **HC220520-91-1** Instrument ID: **MEE-1** Method: **RSK175**

**LCS** Sample ID: **HC220520-91** Units: **UG/L** Analysis Date: **5/23/2022 12:21**  
 Client ID: Run ID: **HC220520-91A** Prep Date: **5/20/2022** DF: **1**

Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
METHANE	154	1	142		108	76-125				25	
ETHANE	296	2	267		111	70-120				25	
PROPANE	440	1	391		112	72-120				25	

**LCSD** Sample ID: **HC220520-91** Units: **UG/L** Analysis Date: **5/23/2022 13:38**  
 Client ID: Run ID: **HC220520-91A** Prep Date: **5/20/2022** DF: **1**

Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
METHANE	160	1	142		112	76-125		154	4	25	
ETHANE	303	2	267		114	70-120		296	2	25	
PROPANE	448	1	391		114	72-120		440	2	25	

**MB** Sample ID: **HC220520-91** Units: **UG/L** Analysis Date: **5/23/2022 12:26**  
 Client ID: Run ID: **HC220520-91A** Prep Date: **5/20/2022** DF: **1**

Analyte	Result	ReportLimit	Qual
METHANE	ND	1	
ETHANE	ND	2	
PROPANE	ND	1	

The following samples were analyzed in this batch:

Client: PDC Energy  
 Work Order: 2205376  
 Project: Postle IC 09-299HNBH

# QC BATCH REPORT

Batch ID: **IP220523-4-3** Instrument ID: **ICP5900** Method: **EPA200.7**

LCS		Sample ID: <b>IP220523-4</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022 09:24</b>				
Client ID:		Run ID: <b>IT220525-1A4</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
CALCIUM	39.9	1	40		100	85-115				20	
MAGNESIUM	38.8	1	40		97	85-115				20	
POTASSIUM	39.8	1	40		100	85-115				20	
SODIUM	40	1	40		100	85-115				20	

LCSD		Sample ID: <b>IP220523-4</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022 09:25</b>				
Client ID:		Run ID: <b>IT220525-1A4</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
CALCIUM	39.8	1	40		100	85-115		39.9	0	20	
MAGNESIUM	38.7	1	40		97	85-115		38.8	0	20	
POTASSIUM	39.7	1	40		99	85-115		39.8	0	20	
SODIUM	39.9	1	40		100	85-115		40	0	20	

MB		Sample ID: <b>IP220523-4</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022 09:22</b>					
Client ID:		Run ID: <b>IT220525-1A4</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>					
Analyte	Result	ReportLimit										Qual
CALCIUM	ND	1										
MAGNESIUM	ND	1										
POTASSIUM	ND	1										
SODIUM	ND	1										

The following samples were analyzed in this batch:

Client: PDC Energy  
 Work Order: 2205376  
 Project: Postle IC 09-299HNBH

# QC BATCH REPORT

Batch ID: **IP220523-4-4** Instrument ID: **ICP5900** Method: **EPA200.7**

LCS		Sample ID: <b>IP220523-4</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022 09:24</b>				
Client ID:		Run ID: <b>IT220525-1A4</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
CALCIUM	39.9	1	40		100	85-115				20	
MAGNESIUM	38.8	1	40		97	85-115				20	
POTASSIUM	39.8	1	40		100	85-115				20	
SODIUM	40	1	40		100	85-115				20	

LCSD		Sample ID: <b>IP220523-4</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022 09:25</b>				
Client ID:		Run ID: <b>IT220525-1A4</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
CALCIUM	39.8	1	40		100	85-115		39.9	0	20	
MAGNESIUM	38.7	1	40		97	85-115		38.8	0	20	
POTASSIUM	39.7	1	40		99	85-115		39.8	0	20	
SODIUM	39.9	1	40		100	85-115		40	0	20	

MB		Sample ID: <b>FP220520-4</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022 09:23</b>				
Client ID:		Run ID: <b>IT220525-1A4</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	Qual								
CALCIUM	ND	1									
MAGNESIUM	ND	1									
POTASSIUM	ND	1									
SODIUM	ND	1									

The following samples were analyzed in this batch:

Client: PDC Energy  
 Work Order: 2205376  
 Project: Postle IC 09-299HNBH

# QC BATCH REPORT

Batch ID: VL220526-3-2 Instrument ID: HPV3 Method: SW8260\_25

LCS		Sample ID: VL220526-3			Units: %REC		Analysis Date: 5/26/2022 16:54				
Client ID:		Run ID: VL220526-3A			Prep Date: 5/26/2022		DF: 1				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
Surr: 4-BROMOFLUOROBENZENE	24.6		25		98	80-120					
Surr: DIBROMOFLUOROMETHANE	26.1		25		104	80-120					
Surr: TOLUENE-D8	23.9		25		96	80-120					
BENZENE	8.86	1	10		89	80-120				20	
TOLUENE	8.57	1	10		86	80-120				20	
ETHYLBENZENE	8.56	1	10		86	80-120				20	
M+P-XYLENE	16.3	1	20		82	80-120				20	
O-XYLENE	8.45	1	10		84	80-120				20	

LCSD		Sample ID: VL220526-3			Units: %REC		Analysis Date: 5/26/2022 17:17				
Client ID:		Run ID: VL220526-3A			Prep Date: 5/26/2022		DF: 1				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
Surr: 4-BROMOFLUOROBENZENE	25.1		25		100	80-120			2		
Surr: DIBROMOFLUOROMETHANE	25.8		25		103	80-120			1		
Surr: TOLUENE-D8	24		25		96	80-120			1		
BENZENE	9.11	1	10		91	80-120		8.86	3	20	
TOLUENE	8.78	1	10		88	80-120		8.57	2	20	
ETHYLBENZENE	8.65	1	10		87	80-120		8.56	1	20	
M+P-XYLENE	16.6	1	20		83	80-120		16.3	2	20	
O-XYLENE	8.59	1	10		86	80-120		8.45	2	20	

MB		Sample ID: VL220526-3			Units: %REC		Analysis Date: 5/26/2022 18:28				
Client ID:		Run ID: VL220526-3A			Prep Date: 5/26/2022		DF: 1				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
Surr: 4-BROMOFLUOROBENZENE	25.3		25		101	80-120					
Surr: DIBROMOFLUOROMETHANE	26.1		25		104	80-120					
Surr: TOLUENE-D8	24.2		25		97	80-120					
BENZENE	ND	1	10								
TOLUENE	ND	1	10								
ETHYLBENZENE	ND	1	10								
M+P-XYLENE	ND	1	20								
O-XYLENE	ND	1	10								
TOTAL XYLENES	ND	1	20								

The following samples were analyzed in this batch:

**Client:** PDC Energy  
**Work Order:** 2205376  
**Project:** Postle IC 09-299HNBH

# QC BATCH REPORT

Batch ID: **AK220526-1-1**      Instrument ID: **Balance**      Method: **SM2320B**

LCS		Sample ID: <b>AK220526-1</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/26/2022</b>				
Client ID:		Run ID: <b>AK220526-1A1</b>			Prep Date: <b>5/26/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
TOTAL ALKALINITY AS CaCO3	102	5	100		102	85-115				15	

MB		Sample ID: <b>AK220526-1</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/26/2022</b>				
Client ID:		Run ID: <b>AK220526-1A1</b>			Prep Date: <b>5/26/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit									
BICARBONATE AS CaCO3	ND	5									
CARBONATE AS CaCO3	ND	5									
TOTAL ALKALINITY AS CaCO3	ND	5									

**The following samples were analyzed in this batch:**

Client: PDC Energy  
 Work Order: 2205376  
 Project: Postle IC 09-299HNBH

# QC BATCH REPORT

Batch ID: **IC220526-2-1** Instrument ID: **IC3** Method: **EPA300.0**

LCS		Sample ID: <b>IC220526-2</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/26/2022 11:14</b>				
Client ID:		Run ID: <b>IC220526-1A1</b>			Prep Date: <b>5/26/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
CHLORIDE	9.92	0.2	10		99	90-110				15	
SULFATE	49.4	1	50		99	90-110				15	

LCSD		Sample ID: <b>IC220526-2</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/26/2022 12:17</b>				
Client ID:		Run ID: <b>IC220526-1A1</b>			Prep Date: <b>5/26/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
CHLORIDE	10	0.2	10		100	90-110		9.92	1	15	
SULFATE	49.6	1	50		99	90-110		49.4	0	15	

MB		Sample ID: <b>IC220526-2</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/26/2022 11:20</b>					
Client ID:		Run ID: <b>IC220526-1A1</b>			Prep Date: <b>5/26/2022</b>		DF: <b>1</b>					
Analyte	Result	ReportLimit										Qual
CHLORIDE	ND	0.2										
SULFATE	ND	1										

The following samples were analyzed in this batch:

**Client:** PDC Energy  
**Work Order:** 2205376  
**Project:** Postle IC 09-299HNBH

## QC BATCH REPORT

Batch ID: **TD220523-1-1**      Instrument ID: **Balance**      Method: **SM2540C**

LCS		Sample ID: <b>TD220523-1</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022</b>				
Client ID:		Run ID: <b>TD220525-1A1</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
TOTAL DISSOLVED SOLIDS	361	20	400		90	85-115				14	

LCSD		Sample ID: <b>TD220523-1</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022</b>				
Client ID:		Run ID: <b>TD220525-1A1</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>				
Analyte	Result	ReportLimit	SPK Val	SPK Ref Value	%REC	Control Limit	Decision Level	RPD Ref Value	RPD	RPD Limit	Qual
TOTAL DISSOLVED SOLIDS	367	20	400		92	85-115		361	2	14	

MB		Sample ID: <b>TD220523-1</b>			Units: <b>MG/L</b>		Analysis Date: <b>5/25/2022</b>					
Client ID:		Run ID: <b>TD220525-1A1</b>			Prep Date: <b>5/23/2022</b>		DF: <b>1</b>					
Analyte	Result	ReportLimit										Qual
TOTAL DISSOLVED SOLIDS	ND	20										

**The following samples were analyzed in this batch:**