

# Session Report

8/18/2022

## Information Panel


Name S004\_BGW020002\_18082022\_114512  
Start Time 8/17/2022 9:45:02 AM  
Stop Time 8/18/2022 9:44:09 AM  
Device Name BGW020002  
Model Type SoundPro DL  
Device Firmware Rev R.13H  
Comments

## Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	53.1 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	FAST	Bandwidth	1	OFF
Leq	2	70 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	FAST			

## Logged Data Chart

S004\_BGW020002\_18082022\_114512: Logged Data Chart



## Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
8/17/2022 9:43:51 AM	Calibration	114.0			
8/18/2022 9:44:46 AM	Verification	114.2			