

Session Report

5/25/2023

Information Panel


Name Rio LC Pad
Start Time 5/23/2023 12:17:51 PM
Stop Time 5/24/2023 7:55:54 AM
Device Name BGW020002
Model Type SoundPro DL
Device Firmware Rev R.13H
Comments

Summary Data Panel

Description	Meter	Value	Description	Meter	Value
Leq	1	52.9 dB			
Exchange Rate	1	3 dB	Weighting	1	A
Response	1	FAST	Bandwidth	1	OFF
Leq	2	75.1 dB			
Exchange Rate	2	3 dB	Weighting	2	C
Response	2	SLOW			

Logged Data Chart

Rio LC Pad: Logged Data Chart



Calibration History

Date	Calibration Action	Level	Cal. Model Type	Serial Number	Cert. Due Date
5/23/2023 12:12:52 PM	Calibration	114.0			
5/24/2023 7:56:37 AM	Verification	114.4			