

Caerus Oil and Gas

Sample Delivery Group: L1619611
Samples Received: 05/24/2023
Project Number:
Description: 909J

Report To: Brett M. , Jake J. , Blair R.
143 Diamond Avenue
Parachute, CO 81635

Entire Report Reviewed By:



Chris Ward
Project Manager

Results relate only to the items tested or calibrated and are reported as rounded values. This test report shall not be reproduced, except in full, without written approval of the laboratory. Where applicable, sampling conducted by Pace Analytical National is performed per guidance provided in laboratory standard operating procedures ENV-SOP-MTJL-0067 and ENV-SOP-MTJL-0068. Where sampling conducted by the customer, results relate to the accuracy of the information provided, and as the samples are received.

Pace Analytical National12065 Lebanon Rd Mount Juliet, TN 37122 615-758-5858 800-767-5859 www.pacenational.com

TABLE OF CONTENTS

Cp: Cover Page	1
Tc: Table of Contents	2
Ss: Sample Summary	3
Cn: Case Narrative	4
Sr: Sample Results	5
20230523-LMSOURCE-(CD29-T) L1619611-01	5
Qc: Quality Control Summary	7
Wet Chemistry by Method 2320 B-2011	7
Wet Chemistry by Method 353.2	9
Wet Chemistry by Method 365.4	10
Wet Chemistry by Method 9040C	11
Wet Chemistry by Method 9050A	12
Wet Chemistry by Method 9056A	13
Metals (ICP) by Method 6010B	15
Volatile Organic Compounds (GC) by Method 8015D/GRO	17
Volatile Organic Compounds (GC/MS) by Method 8260B	18
Semi-Volatile Organic Compounds (GC) by Method 8015M	19
Gl: Glossary of Terms	20
Al: Accreditations & Locations	21
Sc: Sample Chain of Custody	22

¹ Cp

² Tc

³ Ss

⁴ Cn

⁵ Sr

⁶ Qc

⁷ Gl

⁸ Al

⁹ Sc

SAMPLE SUMMARY

20230523-LMSOURCE-(CD29-T) L1619611-01 GW

Collected by: Will Harmon
 Collected date/time: 05/23/23 11:05
 Received date/time: 05/24/23 09:00

Method	Batch	Dilution	Preparation date/time	Analysis date/time	Analyst	Location
Wet Chemistry by Method 2320 B-2011	WG2066739	1	05/26/23 11:01	05/26/23 11:01	ARD	Mt. Juliet, TN
Wet Chemistry by Method 353.2	WG2066027	1	05/24/23 23:44	05/24/23 23:44	AEC	Mt. Juliet, TN
Wet Chemistry by Method 365.4	WG2066829	1	05/25/23 08:47	05/25/23 17:42	UNP	Mt. Juliet, TN
Wet Chemistry by Method 9040C	WG2066099	1	05/26/23 16:10	05/26/23 16:10	MCC	Mt. Juliet, TN
Wet Chemistry by Method 9050A	WG2066539	1	05/27/23 13:31	05/27/23 13:31	NTG	Mt. Juliet, TN
Wet Chemistry by Method 9056A	WG2066157	10	05/25/23 05:20	05/25/23 05:20	MDM	Mt. Juliet, TN
Wet Chemistry by Method 9056A	WG2066157	100	05/25/23 05:33	05/25/23 05:33	MDM	Mt. Juliet, TN
Metals (ICP) by Method 6010B	WG2066606	1	05/30/23 05:47	05/30/23 21:46	SPL	Mt. Juliet, TN
Metals (ICP) by Method 6010B	WG2066606	5	05/30/23 05:47	05/31/23 13:39	SPL	Mt. Juliet, TN
Volatile Organic Compounds (GC) by Method 8015D/GRO	WG2068971	1000	05/31/23 21:53	05/31/23 21:53	ACG	Mt. Juliet, TN
Volatile Organic Compounds (GC/MS) by Method 8260B	WG2067900	1000	05/28/23 19:24	05/28/23 19:24	JAH	Mt. Juliet, TN
Semi-Volatile Organic Compounds (GC) by Method 8015M	WG2068245	1	05/30/23 08:46	05/31/23 20:36	MWS	Mt. Juliet, TN
Semi-Volatile Organic Compounds (GC) by Method 8015M	WG2068245	5	05/30/23 08:46	06/01/23 14:15	MWS	Mt. Juliet, TN

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

CASE NARRATIVE

All sample aliquots were received at the correct temperature, in the proper containers, with the appropriate preservatives, and within method specified holding times, unless qualified or notated within the report. Where applicable, all MDL (LOD) and RDL (LOQ) values reported for environmental samples have been corrected for the dilution factor used in the analysis. All Method and Batch Quality Control are within established criteria except where addressed in this case narrative, a non-conformance form or properly qualified within the sample results. By my digital signature below, I affirm to the best of my knowledge, all problems/anomalies observed by the laboratory as having the potential to affect the quality of the data have been identified by the laboratory, and no information or data have been knowingly withheld that would affect the quality of the data.



Chris Ward
Project Manager

¹ Cp

² Tc

³ Ss

⁴ Cn

⁵ Sr

⁶ Qc

⁷ Gl

⁸ Al

⁹ Sc

Wet Chemistry by Method 2320 B-2011

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
Alkalinity	458		8.45	20.0	1	05/26/2023 11:01	WG2066739
Alkalinity,Bicarbonate	458		8.45	20.0	1	05/26/2023 11:01	WG2066739
Alkalinity,Carbonate	U		8.45	20.0	1	05/26/2023 11:01	WG2066739

Sample Narrative:

L1619611-01 WG2066739: Endpoint pH 4.5 Headspace

Wet Chemistry by Method 353.2

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
Nitrate-Nitrite	U		0.0500	0.100	1	05/24/2023 23:44	WG2066027

Wet Chemistry by Method 365.4

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
Phosphorus>Total	0.913		0.0350	0.100	1	05/25/2023 17:42	WG2066829

Wet Chemistry by Method 9040C

Analyte	Result	Qualifier	Dilution	Analysis	Batch
	pH			date / time	
pH	6.82	<u>T8</u>	1	05/26/2023 16:10	WG2066099

Sample Narrative:

L1619611-01 WG2066099: 6.82 at 20.2C

Wet Chemistry by Method 9050A

Analyte	Result	Qualifier	RDL	Dilution	Analysis	Batch
	umhos/cm		umhos/cm		date / time	
Specific Conductance	18200		10.0	1	05/27/2023 13:31	WG2066539

Sample Narrative:


L1619611-01 WG2066539: at 25C

Wet Chemistry by Method 9056A

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
Bromide	46.3		3.53	10.0	10	05/25/2023 05:20	WG2066157
Chloride	6430		37.9	100	100	05/25/2023 05:33	WG2066157
Fluoride	U		0.640	1.50	10	05/25/2023 05:20	WG2066157
Nitrate as (N)	U		0.480	1.00	10	05/25/2023 05:20	WG2066157
Nitrite as (N)	U		0.420	1.00	10	05/25/2023 05:20	WG2066157
Sulfate	6.73	<u>J</u>	5.94	50.0	10	05/25/2023 05:20	WG2066157

Metals (ICP) by Method 6010B

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
Barium	27.8		0.000736	0.00500	1	05/30/2023 21:46	WG2066606
Boron	6.63		0.0200	0.200	1	05/30/2023 21:46	WG2066606
Calcium	92.8		0.0793	1.00	1	05/30/2023 21:46	WG2066606
Iron	18.5		0.0180	0.100	1	05/30/2023 21:46	WG2066606
Magnesium	7.09		0.0853	1.00	1	05/30/2023 21:46	WG2066606
Manganese	0.339		0.000934	0.0100	1	05/30/2023 21:46	WG2066606
Potassium	26.7		0.261	2.00	1	05/30/2023 21:46	WG2066606



Metals (ICP) by Method 6010B

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
Selenium	0.0145		0.00735	0.0100	1	05/30/2023 21:46	WG2066606
Sodium	3630		2.52	15.0	5	05/31/2023 13:39	WG2066606
Strontium	13.2		0.000640	0.0100	1	05/30/2023 21:46	WG2066606

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Volatile Organic Compounds (GC) by Method 8015D/GRO

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
TPH (GC/FID) Low Fraction	208		31.4	100	1000	05/31/2023 21:53	WG2068971
(S) <i>a,a,a</i> -Trifluorotoluene(FID)	98.4			78.0-120		05/31/2023 21:53	WG2068971

Volatile Organic Compounds (GC/MS) by Method 8260B

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
Benzene	26.7		0.0941	1.00	1000	05/28/2023 19:24	WG2067900
Toluene	36.9		0.278	1.00	1000	05/28/2023 19:24	WG2067900
Ethylbenzene	0.958	J	0.137	1.00	1000	05/28/2023 19:24	WG2067900
Xylenes, Total	12.0		0.174	3.00	1000	05/28/2023 19:24	WG2067900
Naphthalene	U		1.00	5.00	1000	05/28/2023 19:24	WG2067900
(S) Toluene-d8	94.5			80.0-120		05/28/2023 19:24	WG2067900
(S) 4-Bromofluorobenzene	101			77.0-126		05/28/2023 19:24	WG2067900
(S) 1,2-Dichloroethane-d4	110			70.0-130		05/28/2023 19:24	WG2067900

Semi-Volatile Organic Compounds (GC) by Method 8015M

Analyte	Result	Qualifier	MDL	RDL	Dilution	Analysis	Batch
	mg/l		mg/l	mg/l		date / time	
C10-C28 Diesel Range	14.9		0.111	0.500	5	06/01/2023 14:15	WG2068245
C28-C36 Motor Oil Range	1.46		0.0118	0.100	1	05/31/2023 20:36	WG2068245
(S) <i>o</i> -Terphenyl	0.000	J2		52.0-156		05/31/2023 20:36	WG2068245
(S) <i>o</i> -Terphenyl	0.000	J2		52.0-156		06/01/2023 14:15	WG2068245

Sample Narrative:

L1619611-01 WG2068245: Surrogate failure due to matrix interference

Method Blank (MB)

(MB) R3929758-2 05/26/23 09:39

Analyte	MB Result	MB Qualifier	MB MDL	MB RDL
	mg/l		mg/l	mg/l
Alkalinity	U		8.45	20.0
Alkalinity,Bicarbonate	U		8.45	20.0
Alkalinity,Carbonate	U		8.45	20.0

Sample Narrative:

BLANK: Endpoint pH 4.5

L1618948-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1618948-01 05/26/23 10:07 • (DUP) R3929758-3 05/26/23 10:14

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
	mg/l	mg/l		%		%
Alkalinity	28.6	29.3	1	2.20		20
Alkalinity,Bicarbonate	28.6	29.3	1	2.20		20
Alkalinity,Carbonate	U	U	1	0.000		20

Sample Narrative:

OS: Endpoint pH 4.5 Headspace

DUP: Endpoint pH 4.5

L1619875-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1619875-01 05/26/23 11:09 • (DUP) R3929758-4 05/26/23 11:13

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
	mg/l	mg/l		%		%
Alkalinity	131	130	1	0.964		20

Sample Narrative:

OS: Endpoint pH 4.5 Headspace

DUP: Endpoint pH 4.5



Laboratory Control Sample (LCS)

(LCS) R3929758-1 05/26/23 09:30

Analyte	Spike Amount mg/l	LCS Result mg/l	LCS Rec. %	Rec. Limits %	<u>LCS Qualifier</u>
Alkalinity	100	99.4	99.4	90.0-110	

Sample Narrative:

LCS: Endpoint pH 4.5

¹Cp

²Tc

³Ss

⁴Cn

⁵Sr

⁶Qc

⁷Gl

⁸Al

⁹Sc

Method Blank (MB)

(MB) R3929063-1 05/24/23 23:06

Analyte	MB Result	MB Qualifier	MB MDL	MB RDL
Nitrate-Nitrite	U		0.0500	0.100

¹Cp

²Tc

³Ss

⁴Cn

⁵Sr

⁶Qc

⁷Gl

⁸Al

⁹Sc

L1619611-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1619611-01 05/24/23 23:44 • (DUP) R3929063-6 05/24/23 23:45

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Nitrate-Nitrite	U	U	1	0.000		20

L1619500-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1619500-01 05/24/23 23:48 • (DUP) R3929063-8 05/24/23 23:49

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Nitrate-Nitrite	7.80	7.75	5	0.643		20

Laboratory Control Sample (LCS)

(LCS) R3929063-2 05/24/23 23:07

Analyte	Spike Amount	LCS Result	LCS Rec.	Rec. Limits	LCS Qualifier
Nitrate-Nitrite	2.50	2.51	100	90.0-110	

L1619611-01 Original Sample (OS) • Matrix Spike (MS)

(OS) L1619611-01 05/24/23 23:44 • (MS) R3929063-7 05/24/23 23:47

Analyte	Spike Amount	Original Result	MS Result	MS Rec.	Dilution	Rec. Limits	MS Qualifier
Nitrate-Nitrite	2.50	U	2.43	97.2	1	90.0-110	

L1619500-01 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1619500-01 05/24/23 23:48 • (MS) R3929063-9 05/24/23 23:54 • (MSD) R3929063-10 05/24/23 23:56

Analyte	Spike Amount	Original Result	MS Result	MSD Result	MS Rec.	MSD Rec.	Dilution	Rec. Limits	MS Qualifier	MSD Qualifier	RPD	RPD Limits
Nitrate-Nitrite	12.5	7.80	19.8	19.5	95.6	93.2	5	90.0-110			1.53	20

Method Blank (MB)

(MB) R3929542-1 05/25/23 17:11

Analyte	MB Result	MB Qualifier	MB MDL	MB RDL
Phosphorus,Total	U		0.0350	0.100

¹Cp

²Tc

³Ss

⁴Cn

⁵Sr

⁶Qc

⁷Gl

⁸Al

⁹Sc

L1619207-03 Original Sample (OS) • Duplicate (DUP)

(OS) L1619207-03 05/25/23 17:18 • (DUP) R3929542-5 05/25/23 17:19

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Phosphorus,Total	0.452	0.448	1	0.889		20

L1619340-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1619340-01 05/25/23 17:28 • (DUP) R3929542-6 05/25/23 17:29

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Phosphorus,Total	2.25	2.28	1	1.32		20

Laboratory Control Sample (LCS)

(LCS) R3929542-2 05/25/23 17:13

Analyte	Spike Amount	LCS Result	LCS Rec.	Rec. Limits	LCS Qualifier
Phosphorus,Total	3.42	3.20	93.6	83.2-116	

L1619207-02 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1619207-02 05/25/23 17:14 • (MS) R3929542-3 05/25/23 17:15 • (MSD) R3929542-4 05/25/23 17:16

Analyte	Spike Amount	Original Result	MS Result	MSD Result	MS Rec.	MSD Rec.	Dilution	Rec. Limits	MS Qualifier	MSD Qualifier	RPD	RPD Limits
Phosphorus,Total	2.50	0.728	3.08	3.93	94.1	128	1	90.0-110		J3 J5	24.3	20

Sample Narrative:

MSD: Matrix spike failure due to matrix.

L1619611-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1619611-01 05/26/23 16:10 • (DUP) R3929929-2 05/26/23 16:10

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
pH	6.82	6.86	1	0.585		1

Sample Narrative:

OS: 6.82 at 20.2C
DUP: 6.86 at 20.2C

L1620221-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1620221-01 05/26/23 16:10 • (DUP) R3929929-3 05/26/23 16:10

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
pH	4.66	4.65	1	0.215		1

Sample Narrative:

OS: 4.66 at 19.8C
DUP: 4.65 at 19.9C


Laboratory Control Sample (LCS)

(LCS) R3929929-1 05/26/23 16:10

Analyte	Spike Amount	LCS Result	LCS Rec.	Rec. Limits	LCS Qualifier
pH	10.0	10.0	100	99.0-101	

Sample Narrative:

LCS: 10.03 at 19.9C



Method Blank (MB)

(MB) R3930061-1 05/27/23 13:31

Analyte	MB Result	MB Qualifier	MB MDL	MB RDL
Specific Conductance	U		10.0	10.0

Sample Narrative:

BLANK: at 25C

L1617461-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1617461-01 05/27/23 13:31 • (DUP) R3930061-3 05/27/23 13:31

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Specific Conductance	ND	ND	1	0.000		20

Sample Narrative:

OS: at 25C

DUP: at 25C

L1620145-03 Original Sample (OS) • Duplicate (DUP)

(OS) L1620145-03 05/27/23 13:31 • (DUP) R3930061-4 05/27/23 13:31

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
Specific Conductance	733	732	1	0.137		20

Sample Narrative:

OS: at 25C

DUP: at 25C

Laboratory Control Sample (LCS)

(LCS) R3930061-2 05/27/23 13:31

Analyte	Spike Amount	LCS Result	LCS Rec.	Rec. Limits	LCS Qualifier
Specific Conductance	327	326	99.7	85.0-115	

Sample Narrative:

LCS: at 25C

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Method Blank (MB)

(MB) R3931545-1 05/25/23 01:04

Analyte	MB Result	MB Qualifier	MB MDL	MB RDL
	mg/l		mg/l	mg/l
Bromide	U		0.353	1.00
Chloride	U		0.379	1.00
Fluoride	U		0.0640	0.150
Nitrate as (N)	U		0.0480	0.100
Nitrite as (N)	U		0.0420	0.100
Sulfate	U		0.594	5.00

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

L1619469-01 Original Sample (OS) • Duplicate (DUP)

(OS) L1619469-01 05/25/23 01:31 • (DUP) R3931545-3 05/25/23 01:45

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
	mg/l	mg/l		%		%
Bromide	0.472	0.470	1	0.319	U	15
Chloride	7.16	7.11	1	0.678		15
Fluoride	0.246	0.249	1	1.37		15
Nitrate as (N)	0.101	0.111	1	9.93		15
Nitrite as (N)	0.0425	0.0477	1	11.5	U	15
Sulfate	72.5	71.8	1	1.03		15

6 Qc

7 Gl

8 Al

9 Sc

L1617854-05 Original Sample (OS) • Duplicate (DUP)

(OS) L1617854-05 05/25/23 14:26 • (DUP) R3931545-5 05/25/23 14:40

Analyte	Original Result	DUP Result	Dilution	DUP RPD	DUP Qualifier	DUP RPD Limits
	mg/l	mg/l		%		%
Chloride	6.66	6.61	1	0.793		15
Fluoride	0.385	0.388	1	0.776		15
Sulfate	22.4	22.2	1	1.01		15

Laboratory Control Sample (LCS)

(LCS) R3931545-2 05/25/23 01:18

Analyte	Spike Amount	LCS Result	LCS Rec.	Rec. Limits	LCS Qualifier
	mg/l	mg/l	%	%	
Bromide	40.0	40.5	101	80.0-120	
Chloride	40.0	39.2	98.1	80.0-120	
Fluoride	8.00	8.45	106	80.0-120	
Nitrate as (N)	8.00	7.83	97.9	80.0-120	

Laboratory Control Sample (LCS)

(LCS) R3931545-2 05/25/23 01:18

Analyte	Spike Amount mg/l	LCS Result mg/l	LCS Rec. %	Rec. Limits %	<u>LCS Qualifier</u>
Nitrite as (N)	8.00	8.18	102	80.0-120	
Sulfate	40.0	38.6	96.5	80.0-120	

L1619469-01 Original Sample (OS) • Matrix Spike (MS)

(OS) L1619469-01 05/25/23 01:31 • (MS) R3931545-4 05/25/23 01:58

Analyte	Spike Amount mg/l	Original Result mg/l	MS Result mg/l	MS Rec. %	Dilution	Rec. Limits %	<u>MS Qualifier</u>
Bromide	50.0	0.472	51.9	103	1	80.0-120	
Chloride	50.0	7.16	60.0	106	1	80.0-120	
Fluoride	5.00	0.246	5.66	108	1	80.0-120	
Nitrate as (N)	5.00	0.101	5.13	101	1	80.0-120	
Nitrite as (N)	5.00	0.0425	5.35	106	1	80.0-120	
Sulfate	50.0	72.5	122	98.0	1	80.0-120	

L1617854-05 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1617854-05 05/25/23 14:26 • (MS) R3931545-6 05/25/23 14:54 • (MSD) R3931545-7 05/25/23 15:34

Analyte	Spike Amount mg/l	Original Result mg/l	MS Result mg/l	MSD Result mg/l	MS Rec. %	MSD Rec. %	Dilution	Rec. Limits %	<u>MS Qualifier</u>	<u>MSD Qualifier</u>	RPD %	RPD Limits %
Chloride	50.0	6.66	59.2	60.1	105	107	1	80.0-120			1.55	15
Fluoride	5.00	0.385	5.67	5.79	106	108	1	80.0-120			2.06	15
Sulfate	50.0	22.4	73.4	73.5	102	102	1	80.0-120			0.138	15

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Method Blank (MB)

(MB) R3930875-1 05/30/23 21:24

Analyte	MB Result	MB Qualifier	MB MDL	MB RDL
	mg/l		mg/l	mg/l
Barium	U		0.000736	0.00500
Boron	U		0.0200	0.200
Calcium	0.182	U	0.0793	1.00
Iron	U		0.0180	0.100
Magnesium	U		0.0853	1.00
Manganese	U		0.000934	0.0100
Potassium	U		0.261	2.00
Selenium	U		0.00735	0.0100
Sodium	U		0.504	3.00
Strontium	U		0.000640	0.0100

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

Laboratory Control Sample (LCS)

(LCS) R3930875-2 05/30/23 21:26

Analyte	Spike Amount	LCS Result	LCS Rec.	Rec. Limits	LCS Qualifier
	mg/l	mg/l	%	%	
Barium	1.00	1.04	104	80.0-120	
Boron	1.00	1.01	101	80.0-120	
Calcium	10.0	10.5	105	80.0-120	
Iron	10.0	10.1	101	80.0-120	
Magnesium	10.0	10.3	103	80.0-120	
Manganese	1.00	1.02	102	80.0-120	
Potassium	10.0	9.84	98.4	80.0-120	
Selenium	1.00	0.979	97.9	80.0-120	
Sodium	10.0	9.92	99.2	80.0-120	
Strontium	1.00	0.975	97.5	80.0-120	

7 Gl

8 Al

9 Sc

L1619589-01 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1619589-01 05/30/23 21:29 • (MS) R3930875-4 05/30/23 21:35 • (MSD) R3930875-5 05/30/23 21:38

Analyte	Spike Amount	Original Result	MS Result	MSD Result	MS Rec.	MSD Rec.	Dilution	Rec. Limits	MS Qualifier	MSD Qualifier	RPD	RPD Limits
	mg/l	mg/l	mg/l	mg/l	%	%		%			%	%
Barium	1.00	8.03	8.72	8.83	69.0	79.9	1	75.0-125	U		1.25	20
Boron	1.00	7.34	8.06	8.07	72.0	73.8	1	75.0-125	U	U	0.225	20
Calcium	10.0	102	109	109	70.3	69.7	1	75.0-125	U	U	0.0557	20
Iron	10.0	14.7	26.5	26.3	118	117	1	75.0-125			0.720	20
Magnesium	10.0	3.65	13.0	12.7	93.5	90.9	1	75.0-125			1.96	20
Manganese	1.00	1.13	2.07	2.07	93.9	94.6	1	75.0-125			0.351	20
Potassium	10.0	20.6	29.7	29.6	91.4	90.6	1	75.0-125			0.290	20

L1619589-01 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1619589-01 05/30/23 21:29 • (MS) R3930875-4 05/30/23 21:35 • (MSD) R3930875-5 05/30/23 21:38

Analyte	Spike Amount mg/l	Original Result mg/l	MS Result mg/l	MSD Result mg/l	MS Rec. %	MSD Rec. %	Dilution	Rec. Limits %	MS Qualifier	MSD Qualifier	RPD %	RPD Limits %
Selenium	1.00	0.0525	1.44	1.41	139	136	1	75.0-125	<u>J5</u>	<u>J5</u>	2.23	20
Sodium	10.0	3150	3060	3060	0.000	0.000	1	75.0-125	<u>EV</u>	<u>EV</u>	0.170	20
Strontium	1.00	9.39	10.3	9.84	90.6	44.9	1	75.0-125	<u>V</u>	<u>V</u>	4.54	20

¹ Cp

² Tc

³ Ss

⁴ Cn

⁵ Sr

⁶ Qc

⁷ Gl

⁸ Al

⁹ Sc

Method Blank (MB)

(MB) R3931493-2 05/31/23 11:20

Analyte	MB Result mg/l	MB Qualifier	MB MDL mg/l	MB RDL mg/l
TPH (GC/FID) Low Fraction	U		0.0314	0.100
(S) a,a,a-Trifluorotoluene(FID)	99.1			78.0-120

Laboratory Control Sample (LCS) • Laboratory Control Sample Duplicate (LCSD)

(LCS) R3931493-1 05/31/23 10:36 • (LCSD) R3931493-3 05/31/23 23:42

Analyte	Spike Amount mg/l	LCS Result mg/l	LCSD Result mg/l	LCS Rec. %	LCSD Rec. %	Rec. Limits %	LCS Qualifier	LCSD Qualifier	RPD %	RPD Limits %
TPH (GC/FID) Low Fraction	5.50	5.34	4.72	97.1	85.8	72.0-127			12.3	20
(S) a,a,a-Trifluorotoluene(FID)				105	103	78.0-120				

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Method Blank (MB)

(MB) R3931278-3 05/28/23 12:11

Analyte	MB Result mg/l	MB Qualifier	MB MDL mg/l	MB RDL mg/l
Benzene	U		0.0000941	0.00100
Toluene	U		0.000278	0.00100
Ethylbenzene	U		0.000137	0.00100
Xylenes, Total	U		0.000174	0.00300
Naphthalene	U		0.00100	0.00500
(S) Toluene-d8	94.6			80.0-120
(S) 4-Bromofluorobenzene	96.5			77.0-126
(S) 1,2-Dichloroethane-d4	111			70.0-130

Laboratory Control Sample (LCS) • Laboratory Control Sample Duplicate (LCSD)

(LCS) R3931278-1 05/28/23 10:24 • (LCSD) R3931278-2 05/28/23 11:07

Analyte	Spike Amount mg/l	LCS Result mg/l	LCSD Result mg/l	LCS Rec. %	LCSD Rec. %	Rec. Limits %	LCS Qualifier	LCSD Qualifier	RPD %	RPD Limits %
Benzene	0.00500	0.00571	0.00570	114	114	70.0-123			0.175	20
Toluene	0.00500	0.00531	0.00532	106	106	79.0-120			0.188	20
Ethylbenzene	0.00500	0.00538	0.00551	108	110	79.0-123			2.39	20
Xylenes, Total	0.0150	0.0159	0.0160	106	107	79.0-123			0.627	20
Naphthalene	0.00500	0.00532	0.00544	106	109	54.0-135			2.23	20
(S) Toluene-d8				92.6	92.2	80.0-120				
(S) 4-Bromofluorobenzene				100	97.9	77.0-126				
(S) 1,2-Dichloroethane-d4				109	113	70.0-130				

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Method Blank (MB)

(MB) R3931125-1 05/31/23 04:45

Analyte	MB Result mg/l	MB Qualifier	MB MDL mg/l	MB RDL mg/l
C10-C28 Diesel Range	U		0.0222	0.100
C28-C36 Motor Oil Range	U		0.0118	0.100
<i>(S) o-Terphenyl</i>	103			52.0-156

Laboratory Control Sample (LCS) • Laboratory Control Sample Duplicate (LCSD)

(LCS) R3931125-2 05/31/23 05:05 • (LCSD) R3931125-3 05/31/23 05:25

Analyte	Spike Amount mg/l	LCS Result mg/l	LCSD Result mg/l	LCS Rec. %	LCSD Rec. %	Rec. Limits %	LCS Qualifier	LCSD Qualifier	RPD %	RPD Limits %
C10-C28 Diesel Range	1.50	1.46	1.52	97.3	101	50.0-150			4.03	20
<i>(S) o-Terphenyl</i>				106	109	52.0-156				

- 1 Cp
- 2 Tc
- 3 Ss
- 4 Cn
- 5 Sr
- 6 Qc
- 7 Gl
- 8 Al
- 9 Sc

GLOSSARY OF TERMS

Guide to Reading and Understanding Your Laboratory Report


The information below is designed to better explain the various terms used in your report of analytical results from the Laboratory. This is not intended as a comprehensive explanation, and if you have additional questions please contact your project representative.

Results Disclaimer - Information that may be provided by the customer, and contained within this report, include Permit Limits, Project Name, Sample ID, Sample Matrix, Sample Preservation, Field Blanks, Field Spikes, Field Duplicates, On-Site Data, Sampling Collection Dates/Times, and Sampling Location. Results relate to the accuracy of this information provided, and as the samples are received.

Abbreviations and Definitions

MDL	Method Detection Limit.
ND	Not detected at the Reporting Limit (or MDL where applicable).
RDL	Reported Detection Limit.
Rec.	Recovery.
RPD	Relative Percent Difference.
SDG	Sample Delivery Group.
(S)	Surrogate (Surrogate Standard) - Analytes added to every blank, sample, Laboratory Control Sample/Duplicate and Matrix Spike/Duplicate; used to evaluate analytical efficiency by measuring recovery. Surrogates are not expected to be detected in all environmental media.
U	Not detected at the Reporting Limit (or MDL where applicable).
Analyte	The name of the particular compound or analysis performed. Some Analyses and Methods will have multiple analytes reported.
Dilution	If the sample matrix contains an interfering material, the sample preparation volume or weight values differ from the standard, or if concentrations of analytes in the sample are higher than the highest limit of concentration that the laboratory can accurately report, the sample may be diluted for analysis. If a value different than 1 is used in this field, the result reported has already been corrected for this factor.
Limits	These are the target % recovery ranges or % difference value that the laboratory has historically determined as normal for the method and analyte being reported. Successful QC Sample analysis will target all analytes recovered or duplicated within these ranges.
Original Sample	The non-spiked sample in the prep batch used to determine the Relative Percent Difference (RPD) from a quality control sample. The Original Sample may not be included within the reported SDG.
Qualifier	This column provides a letter and/or number designation that corresponds to additional information concerning the result reported. If a Qualifier is present, a definition per Qualifier is provided within the Glossary and Definitions page and potentially a discussion of possible implications of the Qualifier in the Case Narrative if applicable.
Result	The actual analytical final result (corrected for any sample specific characteristics) reported for your sample. If there was no measurable result returned for a specific analyte, the result in this column may state "ND" (Not Detected) or "BDL" (Below Detectable Levels). The information in the results column should always be accompanied by either an MDL (Method Detection Limit) or RDL (Reporting Detection Limit) that defines the lowest value that the laboratory could detect or report for this analyte.
Uncertainty (Radiochemistry)	Confidence level of 2 sigma.
Case Narrative (Cn)	A brief discussion about the included sample results, including a discussion of any non-conformances to protocol observed either at sample receipt by the laboratory from the field or during the analytical process. If present, there will be a section in the Case Narrative to discuss the meaning of any data qualifiers used in the report.
Quality Control Summary (Qc)	This section of the report includes the results of the laboratory quality control analyses required by procedure or analytical methods to assist in evaluating the validity of the results reported for your samples. These analyses are not being performed on your samples typically, but on laboratory generated material.
Sample Chain of Custody (Sc)	This is the document created in the field when your samples were initially collected. This is used to verify the time and date of collection, the person collecting the samples, and the analyses that the laboratory is requested to perform. This chain of custody also documents all persons (excluding commercial shippers) that have had control or possession of the samples from the time of collection until delivery to the laboratory for analysis.
Sample Results (Sr)	This section of your report will provide the results of all testing performed on your samples. These results are provided by sample ID and are separated by the analyses performed on each sample. The header line of each analysis section for each sample will provide the name and method number for the analysis reported.
Sample Summary (Ss)	This section of the Analytical Report defines the specific analyses performed for each sample ID, including the dates and times of preparation and/or analysis.

Qualifier	Description
E	The analyte concentration exceeds the upper limit of the calibration range of the instrument established by the initial calibration (ICAL).
J	The identification of the analyte is acceptable; the reported value is an estimate.
J2	Surrogate recovery limits have been exceeded; values are outside lower control limits.
J3	The associated batch QC was outside the established quality control range for precision.
J5	The sample matrix interfered with the ability to make any accurate determination; spike value is high.
T8	Sample(s) received past/too close to holding time expiration.
V	The sample concentration is too high to evaluate accurate spike recoveries.



ACCREDITATIONS & LOCATIONS

Pace Analytical National 12065 Lebanon Rd Mount Juliet, TN 37122

Alabama	40660	Nebraska	NE-OS-15-05
Alaska	17-026	Nevada	TN000032021-1
Arizona	AZ0612	New Hampshire	2975
Arkansas	88-0469	New Jersey–NELAP	TN002
California	2932	New Mexico ¹	TN00003
Colorado	TN00003	New York	11742
Connecticut	PH-0197	North Carolina	Env375
Florida	E87487	North Carolina ¹	DW21704
Georgia	NELAP	North Carolina ³	41
Georgia ¹	923	North Dakota	R-140
Idaho	TN00003	Ohio–VAP	CL0069
Illinois	200008	Oklahoma	9915
Indiana	C-TN-01	Oregon	TN200002
Iowa	364	Pennsylvania	68-02979
Kansas	E-10277	Rhode Island	LA000356
Kentucky ^{1,6}	KY90010	South Carolina	84004002
Kentucky ²	16	South Dakota	n/a
Louisiana	AI30792	Tennessee ^{1,4}	2006
Louisiana	LA018	Texas	T104704245-20-18
Maine	TN00003	Texas ⁵	LAB0152
Maryland	324	Utah	TN000032021-11
Massachusetts	M-TN003	Vermont	VT2006
Michigan	9958	Virginia	110033
Minnesota	047-999-395	Washington	C847
Mississippi	TN00003	West Virginia	233
Missouri	340	Wisconsin	998093910
Montana	CERT0086	Wyoming	A2LA
A2LA – ISO 17025	1461.01	AIHA-LAP,LLC EMLAP	100789
A2LA – ISO 17025 ⁵	1461.02	DOD	1461.01
Canada	1461.01	USDA	P330-15-00234
EPA–Crypto	TN00003		

¹ Drinking Water ² Underground Storage Tanks ³ Aquatic Toxicity ⁴ Chemical/Microbiological ⁵ Mold ⁶ Wastewater n/a Accreditation not applicable

* Not all certifications held by the laboratory are applicable to the results reported in the attached report.

* Accreditation is only applicable to the test methods specified on each scope of accreditation held by Pace Analytical.

¹ Cp

² Tc

³ Ss

⁴ Cn

⁵ Sr

⁶ Qc

⁷ Gl

⁸ Al

⁹ Sc

1169611

<u>Tracking Numbers</u>		<u>N5A7 Temperature</u>
6126 6537 4048		0.6+0=0.6
6126 6537 4060		3.7+0=3.7