


Site Information	
Site Name	St of Calo Hwy
Date	12/1/2023
Sampler	Kendra (bress)
Weather	27°F cloudy

Gas Meter Information	
Make/Model	AEM 5000
S/N	

Calibration Information	
Gas	Time
Pure N ₂	1309
Fresh Air	1307
50 ppm CO, 10 ppm H ₂ S	1310
15% CH ₄ , 15% CO ₂ , Bal N ₂	1312



Bump Test						
Time		CH ₄ (%)	CO ₂ (%)	O ₂ (%)	H ₂ S (ppm)	CO (ppm)
1316	Calibration Gas	15.0	15.0	0.0	0	0
	Readings	14.8	14.9	0.1	0	0

Field Readings										
Point ID	Time	CH ₄ (%)	CO ₂ (%)	O ₂ (%)	H ₂ S (ppm)	CO (ppm)	Purge Length	Sampled? (Y/N)	Sample ID	Comments
SUP01	1319	0.0	0.3	20.8	0	0	63			
SUP02	1322	0.0	0.3	20.9	0	0	62			
SUP03	1325	0.0	0.4	21.1	0	0	63			
SUP04	1329	0.0	0.5	21.3	0	0	64			
SUP05	1331	0.0	0.4	21.5	0	0	63			

Notes and Observations