


Site Information	
Site Name	Anders 11-17 WH
Date	6/15/2023
Sampler	H. Killian
Weather	62°F cloudy

Gas Meter Information	
Make/Model	GEM5000
S/N	6636

Calibration Information	
Gas	Time
Pure N ₂	1108
Fresh Air	1110
50 ppm CO, 10 ppm H ₂ S	1112
15% CH ₄ , 15% CO ₂ , Bal N ₂	1114

1122
1124
1126
1128



Bump Test						
Time		CH ₄ (%)	CO ₂ (%)	O ₂ (%)	H ₂ S (ppm)	CO (ppm)
1131	Calibration Gas	15.0	15.0	0.0	0.0	0
	Readings	15.1	14.8	0.0	0	11

Field Readings											
Point ID	Time	CH ₄ (%)	CO ₂ (%)	O ₂ (%)	H ₂ S (ppm)	CO (ppm)	Probe Depth (ft)	Sampled? (Y/N)	Sample ID	Comments	
SVPO1	1139	0.1	3.3	16.0	0	4	60	Y	SVPO1	tubing clogged	
SVPO2	1141	0.1	2.5	17.8	0	4	60	Y	SVPO2	tubing clogged	
SVPO3	1146	0.0	2.0	17.5	0	2	61	Y			
SVPO4	1152	0.0	1.7	18.2	0	1	60				
SVPO5	1156	0.0	2.1	17.4	0	2	61				

Notes and Observations