


Soil Vapor Screening Field Form

Site Information	
Site Name	Frank 10-31
Date	1/31/2024
Sampler	H. Mishiki
Weather	Sunny, 60F

Gas Meter Information	
Make/Model	Landtec GEM 5000
S/N	7602

Calibration Information	
Gas	Time
Pure N ₂	10:56
Fresh Air	10:54
50 ppm CO, 10 ppm H ₂ S	10:58
15% CH ₄ , 15% CO ₂ , Bal N ₂	11:00

Site Sketch	
	

Bump Test						
Time		CH ₄ (%)	CO ₂ (%)	O ₂ (%)	H ₂ S (ppm)	CO (ppm)
11:04	Calibration Gas	15.0	15.0	0.0	0	0
	Readings	14.7	14.7	0.3	0	0

Field Readings										
Point ID	Time	CH ₄ (%)	CO ₂ (%)	O ₂ (%)	H ₂ S (ppm)	CO (ppm)	Purge Length	Sampled? (Y/N)	Sample ID	Comments
SVP01	12:41	0.0	0.2	20.8	0	0	65	N	—	
SVP02	12:44	0.0	0.9	20.5	0	0	67	N	—	
SVP03	12:46	0.0	0.6	20.5	0	0	63	N	—	
SVP04	12:48	0.0	0.3	20.6	0	0	62	N	—	
SVP05	—	—	—	—	—	—	—	—	—	Blocked

Notes and Observations

No CH₄ detected → points abandoned