



**ENTHALPY**  
ANALYTICAL

Enthalpy Analytical  
931 West Barkley Ave  
Orange, CA 92868  
(714) 771-6900

enthalpy.com

Lab Job Number : 533037  
Report Level : II  
Report Date : 05/15/2025

**Analytical Report** *prepared for:*

David Watts  
CTEH  
5120 Northshore Drive  
North Little Rock, AR 72118

Project: PROJ-054017 - Bishop Loss of Containment

*Authorized for release by:*

Richard Villafania, Project Manager  
[richard.villafania@enthalpy.com](mailto:richard.villafania@enthalpy.com)

This data package has been reviewed for technical correctness and completeness. Release of this data has been authorized by the Laboratory Manager or the Manager's designee, as verified by the above signature which applies to this PDF file as well as any associated electronic data deliverable files. The results contained in this report meet all requirements of NELAP and pertain only to those samples which were submitted for analysis. This report may be reproduced only in its entirety.

CA ELAP# 1338, NELAP# 4038, SCAQMD LAP# 18LA0518, LACSD ID# 10105, ORELAP# 4197

### Sample Summary

---

David Watts  
CTEH  
5120 Northshore Drive  
North Little Rock, AR  
72118

Lab Job #: 533037  
Project No: PROJ-054017  
Location: Bishop Loss of Containment  
Date Received: 05/14/25

---

<b>Sample ID</b>	<b>Lab ID</b>	<b>Collected</b>	<b>Matrix</b>
GACO513T045-1Q001	533037-001	05/13/25 12:54	Wipe

## Case Narrative

---

CTEH  
5120 Northshore Drive  
North Little Rock, AR 72118  
David Watts

Lab Job Number: 533037  
Project No: PROJ-054017  
Location: Bishop Loss of  
Containment  
Date Received: 05/14/25

---

This data package contains sample and QC results for one wipe sample, requested for the above referenced project on 05/14/25. The sample was received cold and intact.

**TPH-Extractables by GC (EPA 8015M):**

No analytical problems were encountered.

**Volatile Organics by GC/MS (EPA 8260B):**

- High recovery was observed for 1,2,4-trimethylbenzene in the LCS for batch 371345; this analyte was not detected at or above the RL in the associated sample.
- No other analytical problems were encountered.

**Semivolatile Organics by GC/MS SIM (EPA 8270C-SIM):**

No analytical problems were encountered.

**ENTHALPY ANALYTICAL, INC.**  
 931 W. Barkley Ave, Orange, CA 92868  
 Phone: (714) 771-6900 Fax: (714)771-9933  
 Billing: Enthalpy - Orange  
 c/o Montrose Environmental Group  
 P.O. Box 741137, Los Angeles, CA 90074-1137



**Chain of Custody Record**  
 Lab No: 533037  
 Page: 1 of 1  
 Standard: 4 Day: 3 Day: 1 Day: Same Day:

**Turn Around Time (Rush by advanced notice only)**

Matrix: A = Air DW = Drinking Water  
 FL = Food Liquid FS = Food Solid L = Liquid  
 PP = Pure Product S = Solid SeaW = Sea Water  
 SW = Swab W = Water WP = Wipe O = Other

Preservatives: 1 = Na<sub>2</sub>S<sub>2</sub>O<sub>3</sub> 2 = HCl 3 = HNO<sub>3</sub>  
 4 = H<sub>2</sub>SO<sub>4</sub> 5 = NaOH 6 = Other

**CUSTOMER INFORMATION**  
 Company: CTEH  
 Report To: David Watts  
 Email: dwatts@cteh.com  
 Address: 5120 Northshore Dr, North Little Rock, AR 72118  
 Phone: 501-801-8500  
 Fax:

**PROJECT INFORMATION**  
 Name: Bishop Loss of Containment  
 Number: PROJ-054017  
 P.O. #: T045-1  
 Address: Chris Flood  
 Sampled By: Chris Flood

**Analysis Request**  
 EPA 8015B M   
 EPA 8260B   
 EPA 8270C SIM   
 Login 533037

Sample ID	Sampling Date	Sampling Time	Matrix	Container No. / Size	Pres.
1 GACO0513T045-1Q001	5-13-25	1254	WP	3	6
2					
3					
4					
5					
6					
7					
8					
9					
10					

South front wall siding

**Signature**  
 Relinquished By: *Chris Flood*  
 Received By: *Chris Flood*  
 Relinquished By:  
 Received By:

**Print Name**  
 Chris Flood  
 FLOOD

**Company / Title**  
 CTEH / ES  
 ES

**Date / Time**  
 5-13-25 / 1615  
 5/14/25 0740

S.C.C. PL 13 C.F.S. 6.0

## SAMPLE RECEIPT CHECKLIST


**Section 1: General Info**

 Date Received: 05/14/25 WO# 533037 Client: CTEH
**Section 2: Shipping / Custody**

 Are custody seals present?  Yes  No

 Custody seals intact on arrival?  N/A  Yes  No  On cooler / box  On samples  
 Courier  Walk-In  Field Sampling  Shipping Info: FedEx
**Section 3a: Condition / Packaging**
 Outside 0.0 - 6.0°C (0.0 - 10.0°C for microbiology) (PM notified)

 Date Opened 05/14/25 By (initials) FPD Type of ice used:  Wet  Blue/Gel  None

 Samples received on ice directly from the field; cooling process had begun. (if checked, skip temperatures)

 Sample matrix doesn't require cooling (e.g. air, bulk PCB). (if checked, skip temperatures)

 If no cooler: Observed/Adjusted Temp (°C): \_\_\_\_\_ / \_\_\_\_\_ Thermometer/IR Gun: IR13 CF: 0.0

 Cooler Temp (°C) #1: 5.6 / 5.6 #2: \_\_\_\_\_ / \_\_\_\_\_ #3: \_\_\_\_\_ / \_\_\_\_\_ #4: \_\_\_\_\_ / \_\_\_\_\_ #5: \_\_\_\_\_ / \_\_\_\_\_ #6: \_\_\_\_\_ / \_\_\_\_\_

**Section 3b: Microbiology Samples**
 No microbiology samples submitted (skip 3b)

 Within temp range 0.0 - 10.0°C or received on ice directly from field.

 Adequate headspace for microbiology analysis.

**Section 3c: Air Samples**
 No air samples submitted (skip 3c)

 1.4L Canisters  6L Canisters  Tedlar Bags  MCE Cassettes  Sorbent Tubes  Other \_\_\_\_\_

**Section 4: Containers / Labels / Samples**

	YES	NO	N/A
1) Were custody papers present, filled properly, and legible?	x		
2) Is the sampler's name present on the CoC?	x		
3) Were containers received in good condition (unbroken / unopened / uncompromised)?	x		
4) Were the samples bagged? (required for microbiology samples; recommended for soil samples)	x		
5) Were all of, and only, the correct samples received?	x		
6) Are sample labels present, legible, and in agreement with the CoC?		x	
7) Does the container count match the CoC?	x		
8) Was sufficient sample volume / mass received for the analyses requested?	x		
9) Were samples received in proper containers for the analyses requested?	x		
10) Were samples received with > 1/2 holding time remaining?	x		
11) Are samples properly preserved as indicated by CoC / labels?	x		
12) Unpreserved VOAs received - If necessary, was the hold time changed in LIMS?			x
13) Are VOA vials free from headspace/bubbles > 6mm?			x

**Section 5: Explanations / Comments**

(If no comments are made, then no discrepancies noted.)

*4.6 - SAMPLING DATE DISCREPANCY - 5/13/25 PER CoC, 5/12/25 PER LABEL*

 No additional discrepancies

Date Logged	<u>05/13/25</u>	By (print)	<u>GCK</u>	(sign)	<i>[Signature]</i>
Date Labeled	<u>05/14/25</u>	By (print)	<u>FPD</u>	(sign)	<i>[Signature]</i>

ORIGIN ID:GXVA (501) 549-9862  
PROJ - 054017  
CTR/TOXOCODOLOGY ENVIRO HLTH  
5120 NORTSHORE DR

NORTH LITTLE ROCK, AR 72118  
UNITED STATES US

SHIP DATE: 13MAY25  
ACTWGT: 22.55 LB  
CAD: 6995978/SSF02600  
DIMS: 25x13x14 IN

BILL THIRD PARTY

Brt # 15629 3039 009020EXP 02/26

TO **ENTHALPY ANALYTICAL INC**

**931 W BARKLEY AVE**


**ORANGE CA 92868**

(714) 771-6900

REF:

YRU:

DEPT:



**FedEx**  
Express



REL#  
3785346


TRK# 2886 2847 0960  
0201

**WED - 14 MAY 8:00A**  
**FIRST OVERNIGHT**

**X1 APVA**

**AHS**  
**92868**

CA-US **SNA**



## Analysis Results for 533037

David Watts  
 CTEH  
 5120 Northshore Drive  
 North Little Rock, AR 72118

Lab Job #: 533037  
 Project No: PROJ-054017  
 Location: Bishop Loss of Containment  
 Date Received: 05/14/25

**Sample ID: GACO0513T045-1Q001      Lab ID: 533037-001      Collected: 05/13/25 12:54**  
**Matrix: Wipe**

533037-001 Analyte	Result	Qual	Units	RL	DF	Batch	Prepared	Analyzed	Chemist
Method: EPA 8015M Prep Method: EPA 3580M									
GRO C6-C10	ND		mg/s	1.0	10	371279	05/14/25	05/14/25	DIB
DRO C10-C28	ND		mg/s	1.0	10	371279	05/14/25	05/14/25	DIB
ORO C28-C44	ND		mg/s	1.0	10	371279	05/14/25	05/14/25	DIB
<b>Surrogates</b>				<b>Limits</b>					
n-Triacontane	112%		%REC	70-130	10	371279	05/14/25	05/14/25	DIB
Method: EPA 8260B Prep Method: EPA 5030B									
Benzene	ND		ug/s	1.3	50	371345	05/14/25	05/14/25	ZST
Toluene	ND		ug/s	1.3	50	371345	05/14/25	05/14/25	ZST
Ethylbenzene	ND		ug/s	1.3	50	371345	05/14/25	05/14/25	ZST
m,p-Xylenes	ND		ug/s	1.3	50	371345	05/14/25	05/14/25	ZST
o-Xylene	ND		ug/s	1.3	50	371345	05/14/25	05/14/25	ZST
1,3,5-Trimethylbenzene	ND		ug/s	1.3	50	371345	05/14/25	05/14/25	ZST
1,2,4-Trimethylbenzene	ND		ug/s	1.3	50	371345	05/14/25	05/14/25	ZST
Xylene (total)	ND		ug/s	2.5	50	371345	05/14/25	05/14/25	ZST
<b>Surrogates</b>				<b>Limits</b>					
Dibromofluoromethane	93%		%REC	70-130	50	371345	05/14/25	05/14/25	ZST
1,2-Dichloroethane-d4	111%		%REC	70-130	50	371345	05/14/25	05/14/25	ZST
Toluene-d8	102%		%REC	70-130	50	371345	05/14/25	05/14/25	ZST
Bromofluorobenzene	101%		%REC	70-130	50	371345	05/14/25	05/14/25	ZST
Method: EPA 8270C-SIM Prep Method: EPA 3580M									
Acenaphthene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Anthracene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Benzo(a)anthracene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Benzo(b)fluoranthene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Benzo(k)fluoranthene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Benzo(a)pyrene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Chrysene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Dibenz(a,h)anthracene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Fluoranthene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Fluorene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Indeno(1,2,3-cd)pyrene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
1-Methylnaphthalene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
2-Methylnaphthalene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Naphthalene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
Pyrene	ND		ug/s	2.0	10	371162	05/14/25	05/14/25	TJW
<b>Surrogates</b>				<b>Limits</b>					
Nitrobenzene-d5	100%		%REC	40-120	10	371162	05/14/25	05/14/25	TJW
2-Fluorobiphenyl	96%		%REC	46-120	10	371162	05/14/25	05/14/25	TJW
Terphenyl-d14	114%		%REC	43-120	10	371162	05/14/25	05/14/25	TJW

## Analysis Results for 533037

533037-001 Analyte	Result	Qual	Units	RL	DF	Batch	Prepared	Analyzed	Chemist
--------------------	--------	------	-------	----	----	-------	----------	----------	---------

ND Not Detected

### Batch QC

<b>Type: Blank</b>	<b>Lab ID: QC1257178</b>	<b>Batch: 371279</b>
<b>Matrix: Wipe</b>	<b>Method: EPA 8015M</b>	<b>Prep Method: EPA 3580M</b>

QC1257178 Analyte	Result	Qual	Units	RL	Prepared	Analyzed
GRO C6-C10	ND		mg/s	1.0	05/14/25	05/14/25
DRO C10-C28	ND		mg/s	1.0	05/14/25	05/14/25
ORO C28-C44	ND		mg/s	1.0	05/14/25	05/14/25
Surrogates				Limits		
n-Triacontane	106%		%REC	70-130	05/14/25	05/14/25

<b>Type: Lab Control Sample</b>	<b>Lab ID: QC1257179</b>	<b>Batch: 371279</b>
<b>Matrix: Wipe</b>	<b>Method: EPA 8015M</b>	<b>Prep Method: EPA 3580M</b>

QC1257179 Analyte	Result	Spiked	Units	Recovery	Qual	Limits
Diesel C10-C28	4.336	5.000	mg/s	87%		76-122
Surrogates						
n-Triacontane	0.2087	0.2000	mg/s	104%		70-130

<b>Type: Lab Control Sample Duplicate</b>	<b>Lab ID: QC1257180</b>	<b>Batch: 371279</b>
<b>Matrix: Wipe</b>	<b>Method: EPA 8015M</b>	<b>Prep Method: EPA 3580M</b>

QC1257180 Analyte	Result	Spiked	Units	Recovery	Qual	Limits	RPD	RPD Lim
Diesel C10-C28	4.373	5.000	mg/s	87%		76-122	1	20
Surrogates								
n-Triacontane	0.2124	0.2000	mg/s	106%		70-130		

<b>Type: Lab Control Sample</b>	<b>Lab ID: QC1257348</b>	<b>Batch: 371345</b>
<b>Matrix: Wipe</b>	<b>Method: EPA 8260B</b>	<b>Prep Method: EPA 5030B</b>

QC1257348 Analyte	Result	Spiked	Units	Recovery	Qual	Limits
Benzene	0.2760	0.2500	ug/s	110%		70-130
Toluene	0.2746	0.2500	ug/s	110%		70-130
Ethylbenzene	0.2817	0.2500	ug/s	113%		80-120
m,p-Xylenes	0.5596	0.5000	ug/s	112%		80-123
o-Xylene	0.2838	0.2500	ug/s	114%		80-120
1,3,5-Trimethylbenzene	0.3196	0.2500	ug/s	128%		80-132
1,2,4-Trimethylbenzene	0.3151	0.2500	ug/s	126%	*	80-125
Surrogates						
Dibromofluoromethane	0.2533	0.2500	ug/s	101%		70-130
1,2-Dichloroethane-d4	0.2724	0.2500	ug/s	109%		70-130
Toluene-d8	0.2570	0.2500	ug/s	103%		70-130
Bromofluorobenzene	0.2582	0.2500	ug/s	103%		70-130

## Batch QC

<b>Type: Blank</b>	<b>Lab ID: QC1257349</b>	<b>Batch: 371345</b>
<b>Matrix: Wipe</b>	<b>Method: EPA 8260B</b>	<b>Prep Method: EPA 5030B</b>

QC1257349 Analyte	Result	Qual	Units	RL	Prepared	Analyzed
Benzene	ND		ug/s	1.3	05/14/25	05/14/25
Toluene	ND		ug/s	1.3	05/14/25	05/14/25
Ethylbenzene	ND		ug/s	1.3	05/14/25	05/14/25
m,p-Xylenes	ND		ug/s	1.3	05/14/25	05/14/25
o-Xylene	ND		ug/s	1.3	05/14/25	05/14/25
1,3,5-Trimethylbenzene	ND		ug/s	1.3	05/14/25	05/14/25
1,2,4-Trimethylbenzene	ND		ug/s	1.3	05/14/25	05/14/25
Xylene (total)	ND		ug/s	2.5	05/14/25	05/14/25
Surrogates	Limits					
Dibromofluoromethane	91%		%REC	70-130	05/14/25	05/14/25
1,2-Dichloroethane-d4	111%		%REC	70-130	05/14/25	05/14/25
Toluene-d8	103%		%REC	70-130	05/14/25	05/14/25
Bromofluorobenzene	100%		%REC	70-130	05/14/25	05/14/25

<b>Type: Matrix Spike</b>	<b>Lab ID: QC1257350</b>	<b>Batch: 371345</b>
<b>Matrix (Source ID): Wipe (533038-001)</b>	<b>Method: EPA 8260B</b>	<b>Prep Method: EPA 5030B</b>

QC1257350 Analyte	Result	Source Sample Result	Spiked	Units	Recovery	Qual	Limits	DF
Benzene	5.542	ND	5.000	ug/s	111%		60-140	50
Toluene	5.613	ND	5.000	ug/s	112%		60-140	50
Ethylbenzene	5.730	ND	5.000	ug/s	115%		70-130	50
m,p-Xylenes	11.37	ND	10.00	ug/s	114%		70-133	50
o-Xylene	5.660	ND	5.000	ug/s	113%		70-130	50
1,3,5-Trimethylbenzene	6.336	ND	5.000	ug/s	127%		70-142	50
1,2,4-Trimethylbenzene	6.244	ND	5.000	ug/s	125%		70-135	50
Surrogates								
Dibromofluoromethane	12.24		12.50	ug/s	98%		70-130	50
1,2-Dichloroethane-d4	13.75		12.50	ug/s	110%		70-130	50
Toluene-d8	12.76		12.50	ug/s	102%		70-130	50
Bromofluorobenzene	12.72		12.50	ug/s	102%		70-130	50

### Batch QC

<b>Type: Matrix Spike Duplicate</b>	<b>Lab ID: QC1257351</b>	<b>Batch: 371345</b>
<b>Matrix (Source ID): Wipe (533038-001)</b>	<b>Method: EPA 8260B</b>	<b>Prep Method: EPA 5030B</b>

QC1257351 Analyte	Result	Source Sample Result	Spiked	Units	Recovery	Qual	Limits	RPD	RPD Lim	DF
Benzene	5.742	ND	5.000	ug/s	115%		60-140	4	30	50
Toluene	5.745	ND	5.000	ug/s	115%		60-140	2	30	50
Ethylbenzene	5.852	ND	5.000	ug/s	117%		70-130	2	30	50
m,p-Xylenes	11.60	ND	10.00	ug/s	116%		70-133	2	30	50
o-Xylene	5.843	ND	5.000	ug/s	117%		70-130	3	30	50
1,3,5-Trimethylbenzene	6.596	ND	5.000	ug/s	132%		70-142	4	39	50
1,2,4-Trimethylbenzene	6.436	ND	5.000	ug/s	129%		70-135	3	30	50
<b>Surrogates</b>										
Dibromofluoromethane	12.01		12.50	ug/s	96%		70-130			50
1,2-Dichloroethane-d4	13.51		12.50	ug/s	108%		70-130			50
Toluene-d8	12.76		12.50	ug/s	102%		70-130			50
Bromofluorobenzene	12.58		12.50	ug/s	101%		70-130			50

<b>Type: Blank</b>	<b>Lab ID: QC1256794</b>	<b>Batch: 371162</b>
<b>Matrix: Wipe</b>	<b>Method: EPA 8270C-SIM</b>	<b>Prep Method: EPA 3580M</b>

QC1256794 Analyte	Result	Qual	Units	RL	Prepared	Analyzed
Acenaphthene	ND		ug/s	2.0	05/13/25	05/13/25
Anthracene	ND		ug/s	2.0	05/13/25	05/13/25
Benzo(a)anthracene	ND		ug/s	2.0	05/13/25	05/13/25
Benzo(b)fluoranthene	ND		ug/s	2.0	05/13/25	05/13/25
Benzo(k)fluoranthene	ND		ug/s	2.0	05/13/25	05/13/25
Benzo(a)pyrene	ND		ug/s	2.0	05/13/25	05/13/25
Chrysene	ND		ug/s	2.0	05/13/25	05/13/25
Dibenz(a,h)anthracene	ND		ug/s	2.0	05/13/25	05/13/25
Fluoranthene	ND		ug/s	2.0	05/13/25	05/13/25
Fluorene	ND		ug/s	2.0	05/13/25	05/13/25
Indeno(1,2,3-cd)pyrene	ND		ug/s	2.0	05/13/25	05/13/25
1-Methylnaphthalene	ND		ug/s	2.0	05/13/25	05/13/25
2-Methylnaphthalene	ND		ug/s	2.0	05/13/25	05/13/25
Naphthalene	ND		ug/s	2.0	05/13/25	05/13/25
Pyrene	ND		ug/s	2.0	05/13/25	05/13/25
<b>Surrogates</b>				<b>Limits</b>		
Nitrobenzene-d5	86%		%REC	40-120	05/13/25	05/13/25
2-Fluorobiphenyl	84%		%REC	46-120	05/13/25	05/13/25
Terphenyl-d14	97%		%REC	43-120	05/13/25	05/13/25

## Batch QC

<b>Type:</b> Lab Control Sample	<b>Lab ID:</b> QC1256795	<b>Batch:</b> 371162
<b>Matrix:</b> Wipe	<b>Method:</b> EPA 8270C-SIM	<b>Prep Method:</b> EPA 3580M

QC1256795 Analyte	Result	Spiked	Units	Recovery	Qual	Limits
Acenaphthene	6.411	8.000	ug/s	80%		49-120
Anthracene	7.081	8.000	ug/s	89%		46-120
Benzo(a)anthracene	6.754	8.000	ug/s	84%		52-120
Benzo(b)fluoranthene	6.343	8.000	ug/s	79%		51-120
Benzo(k)fluoranthene	6.402	8.000	ug/s	80%		51-120
Benzo(a)pyrene	6.196	8.000	ug/s	77%		47-120
Chrysene	6.401	8.000	ug/s	80%		52-120
Dibenz(a,h)anthracene	6.358	8.000	ug/s	79%		41-120
Fluoranthene	6.713	8.000	ug/s	84%		50-120
Fluorene	6.844	8.000	ug/s	86%		51-120
Indeno(1,2,3-cd)pyrene	6.295	8.000	ug/s	79%		43-120
1-Methylnaphthalene	7.808	8.000	ug/s	98%		52-120
2-Methylnaphthalene	7.401	8.000	ug/s	93%		53-120
Naphthalene	6.953	8.000	ug/s	87%		50-120
Pyrene	6.651	8.000	ug/s	83%		48-120
<b>Surrogates</b>						
Nitrobenzene-d5	7.316	8.000	ug/s	91%		40-120
2-Fluorobiphenyl	7.894	8.000	ug/s	99%		46-120
Terphenyl-d14	8.069	8.000	ug/s	101%		43-120

<b>Type:</b> Lab Control Sample Duplicate	<b>Lab ID:</b> QC1256796	<b>Batch:</b> 371162
<b>Matrix:</b> Wipe	<b>Method:</b> EPA 8270C-SIM	<b>Prep Method:</b> EPA 3580M

QC1256796 Analyte	Result	Spiked	Units	Recovery	Qual	Limits	RPD	RPD Lim
Acenaphthene	5.509	8.000	ug/s	69%		49-120	15	20
Anthracene	5.970	8.000	ug/s	75%		46-120	17	20
Benzo(a)anthracene	5.830	8.000	ug/s	73%		52-120	15	20
Benzo(b)fluoranthene	5.478	8.000	ug/s	68%		51-120	15	20
Benzo(k)fluoranthene	5.421	8.000	ug/s	68%		51-120	17	20
Benzo(a)pyrene	5.293	8.000	ug/s	66%		47-120	16	20
Chrysene	5.528	8.000	ug/s	69%		52-120	15	20
Dibenz(a,h)anthracene	5.466	8.000	ug/s	68%		41-120	15	20
Fluoranthene	5.751	8.000	ug/s	72%		50-120	15	20
Fluorene	5.884	8.000	ug/s	74%		51-120	15	20
Indeno(1,2,3-cd)pyrene	5.392	8.000	ug/s	67%		43-120	15	20
1-Methylnaphthalene	6.607	8.000	ug/s	83%		52-120	17	20
2-Methylnaphthalene	6.244	8.000	ug/s	78%		53-120	17	20
Naphthalene	5.959	8.000	ug/s	74%		50-120	15	20
Pyrene	5.744	8.000	ug/s	72%		48-120	15	20
<b>Surrogates</b>								
Nitrobenzene-d5	7.112	8.000	ug/s	89%		40-120		
2-Fluorobiphenyl	7.570	8.000	ug/s	95%		46-120		
Terphenyl-d14	7.914	8.000	ug/s	99%		43-120		

\* Value is outside QC limits  
ND Not Detected