

Calibration Report

Report Date **12/12/25 10:37 AM**

Collection Time

12/12/25 10:35 AM

Section 1 - Meter

Device ID **100109348**
 Location **Cowboy Carl 31-6**

System **Contact # 719-846-7898**
 Field **Raton Basin**
 State **Colorado**
 Producer **Evergreen Natural Resources**
 Purchaser **Colorado Interstate Gas**

Section 2 - Setup

Current Plate Size **0.5**
 Pipe Size **2.067**
 Atmos. Pressure **11.3**
 Pressure Base **14.73**
 Tap Type **Flange**
 SP Tap Location **Up Stream**
 Temperature Bias **0**
 Temperature Base **60**
 Specific Gravity **0.56**

Plate Inspected? Yes No
 Edges Sharp? Yes No
 Seal Ring Damaged? Yes No
 Dirty? Yes No
 Plate Mic'd? Yes No
 Pitted? Yes No
 Nicked? Yes No
 Warped? Yes No
 Bevel Downstream? Yes No

Section 3 - Meter Tube Inspection

Tube Inspected? Yes No
 Dirty? Yes No
 Pitted? Yes No
 Nicked? Yes No

Section 4 - Calibration Data

DIFFERENTIAL				STATIC PRESSURE				TEMPERATURE			
FOUND		LEFT		FOUND		LEFT		FOUND		LEFT	
Test	Meter	Test	Meter	Test	Meter	Test	Meter	Test	Meter	Test	Meter
0.00	0.00	0.00	0.00	11.30	11.28	11.30	11.28				
1.00	0.99	1.00	0.99	13.80	13.85	13.80	13.85				
100.00	99.98	100.00	99.98	111.30	111.20	111.30	111.20				
75.00	75.00	75.00	75.00	61.30	61.27	61.30	61.27				
50.00	50.04	50.00	50.04								
25.00	25.01	25.00	25.01								

DP check with pressure on SP

FOUND		LEFT	
DP	SP	DP	SP

Cal. Range		
DP	SP	Temp.
0	0	0
100	114	150

Gas Sample Taken Yes No
 H2S analysis done Yes No
 PPM or %
 H2O Analysis Done Yes No

REMARKS

Signature / Company

Calibrated by Dylan Hamman	Date
Witness	12/12/2025